

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010515\
 Data File : BE089324.D
 Acq On : 5 Jan 2015 20:57
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 06 14:12:11 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIM-BE121914.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 20 00:43:57 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	124	0.00
2	n-Nitrosodimethylamine	1.000	1.238	-23.8	152	0.00
3 S	2-Fluorophenol	1.000	1.232	-23.2	146	0.05
4 S	Phenol-d6	1.000	1.080	-8.0	128	0.07
5 C	Phenol	1.000	1.133	-13.3	134	0.07
6	bis(2-Chloroethyl)ether	1.000	1.148	-14.8	138	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	121	0.00
8 S	Nitrobenzene-d5	1.000	1.034	-3.4	125	0.00
9	Nitrobenzene	1.000	1.151	-15.1	138	0.00
10	2,4-Dimethylphenol	1.000	1.195	-19.5	143	0.04
11 C	2,4-Dichlorophenol	1.000	1.195	-19.5	143	0.03
12	Hexachlorobutadiene	1.000	1.085	-8.5	130	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	1.000	1.117	-11.7	131	0.00
15 S	2-Fluorobiphenyl	1.000	1.021	-2.1	117	0.00
16 P	2,4-Dinitrophenol	1.000	0.750	25.0	174	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	120	0.00
18	Hexachlorobenzene	1.000	1.115	-11.5	130	0.00
19 C	Pentachlorophenol	1.000	1.386	-38.6#	209	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	130	0.00
21	Benzidine	1.000	0.933	6.7	129	0.00
22 S	Terphenyl-d14	1.000	1.005	-0.5	130	0.00
23	Benzo(a)anthracene	1.000	1.193	-19.3	159	0.00
24	3,3'-Dichlorobenzidine	1.000	1.344	-34.4#	177	0.00
25	Chrysene	1.000	1.124	-12.4	144	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	1.369	-36.9#	196	0.00
27 I	Perylene-d12	5.000	5.000	0.0	138	0.00
28	Benzo(b)fluoranthene	1.000	1.333	-33.3#	183	0.00
29	Benzo(k)fluoranthene	1.000	1.120	-12.0	151	0.00
30 C	Benzo(a)pyrene	1.000	1.343	-34.3#	190	0.00
31	Dibenzo(a,h)anthracene	1.000	1.385	-38.5#	195	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 2