




Data Path : Z:\HPCHEM1\BNA_E\Data\BE010617\
Data File : BE092629.D
Acq On : 6 Jan 2017 13:08
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 07 06:01:40 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 06 15:43:51 2017
Response via : Initial Calibration



TIC: BE092629.D

(69) Pentachlorophenol (C)




10.908min (-0.066) 49.99ng

response 12

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	34.48#
264.00	62.90	24.14#
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE010617\
Data File : BE092629.D
Acq On : 6 Jan 2017 13:08
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 07 06:01:40 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 06 15:43:51 2017
Response via : Initial Calibration



TIC: BE092629.D

(76) Benzidine

12.447min (-0.046) -20.00ng

response 147

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	168.64#
183.00	11.50	53.25#
0.00	0.00	0.00