

Data Path : Z:\HPCHEM1\BNA E\Data\BE010617\
 Data File : BE092630.D
 Acq On : 6 Jan 2017 13:47
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 07 03:31:01 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	114	0.00
2	1,4-Dioxane	0.400	0.224	44.0#	88	0.00
3	n-Nitrosodimethylamine	0.400	0.401	-0.3	107	0.01
4 S	2-Fluorophenol	0.400	0.403	-0.8	110	0.00
5 S	Phenol-d6	0.400	0.421	-5.2	125	0.00
6	bis(2-Chloroethyl)ether	0.400	0.418	-4.5	145	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.430	-7.5	113	0.00
9	Naphthalene	0.400	0.380	5.0	114	0.00
10	Hexachlorobutadiene	0.400	0.381	4.8	113	0.00
11	2-Methylnaphthalene	0.400	0.352	12.0	110	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	107	0.00
13 S	2,4,6-Tribromophenol	0.400	0.416	-4.0	107	0.00
14 S	2-Fluorobiphenyl	0.400	0.370	7.5	109	0.00
15	Acenaphthylene	0.400	0.357	10.8	108	0.00
16	Acenaphthene	0.400	0.368	8.0	107	0.00
17	Fluorene	0.400	0.367	8.3	109	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	109	0.00
19	Hexachlorobenzene	0.400	0.360	10.0	106	0.00
20	Pentachlorophenol	0.400	0.413	-3.2	126	0.00
21	Phenanthrene	0.400	0.351	12.3	109	0.00
22	Anthracene	0.400	0.345	13.8	111	0.00
23	Fluoranthene	0.400	0.352	12.0	110	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	112	0.00
25	Pyrene	0.400	0.338	15.5	108	0.00
26 S	Terphenyl-d14	0.400	0.349	12.8	110	0.00
27	Benzo(a)anthracene	0.400	0.350	12.5	114	0.00
28	Chrysene	0.400	0.358	10.5	109	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.318	20.5	93	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.322	19.5	105	0.00
31 I	Perylene-d12	5.000	5.000	0.0	108	0.01
32	Benzo(b)fluoranthene	0.400	0.338	15.5	104	0.00
33	Benzo(k)fluoranthene	0.400	0.376	6.0	117	0.00
34 C	Benzo(a)pyrene	0.400	0.342	14.5	110	0.01
35	Dibenzo(a,h)anthracene	0.400	0.321	19.8	103	0.02
36	Benzo(g,h,i)perylene	0.400	0.345	13.8	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0