

Data Path : Z:\HPCHEM1\BNA E\DATA\BE010617\
 Data File : BE092630.D
 Acq On : 6 Jan 2017 13:47
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

Sohil
 1/9/2017 10:46:21 AM

Quant Time: Jan 07 03:31:01 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	8814	5.00	ng	0.00
7) Naphthalene-d8	10.93	136	42518	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	29187	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	63859	5.00	ng	0.00
24) Chrysene-d12	21.72	240	85369	5.00	ng	0.00
31) Perylene-d12	24.08	264	74587	5.00	ng	0.01

System Monitoring Compounds

4) 2-Fluorophenol	5.69	112	445	0.40	ng	0.00
5) Phenol-d6	7.38	99	662	0.42	ng	0.00
8) Nitrobenzene-d5	9.36	82	764m	0.43	ng	0.00
13) 2,4,6-Tribromophenol	16.31	330	209	0.42	ng	0.00
14) 2-Fluorobiphenyl	13.43	172	2581	0.37	ng	0.00
26) Terphenyl-d14	20.17	244	3650	0.35	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.39	88	393m	0.22	ng	
3) n-Nitrosodimethylamine	3.80	42	312	0.40	ng	# 91
6) bis(2-Chloroethyl)ether	7.59	93	810m	0.42	ng	
9) Naphthalene	10.97	128	3314	0.38	ng	97
10) Hexachlorobutadiene	11.26	225	505	0.38	ng	98
11) 2-Methylnaphthalene	12.64	142	1952	0.35	ng	98
15) Acenaphthylene	14.51	152	14395	0.36	ng	100
16) Acenaphthene	14.85	154	2444	0.37	ng	96
17) Fluorene	15.86	166	2996	0.37	ng	99
19) Hexachlorobenzene	16.87	284	981m	0.36	ng	
20) Pentachlorophenol	17.23	266	147	0.41	ng	92
21) Phenanthrene	17.58	178	5652	0.35	ng	# 95
22) Anthracene	17.69	178	5073	0.34	ng	99
23) Fluoranthene	19.59	202	24513	0.35	ng	100
25) Pyrene	19.95	202	25383	0.34	ng	100
27) Benzo(a)anthracene	21.70	228	26814m	0.35	ng	
28) Chrysene	21.75	228	31831m	0.36	ng	
29) Bis(2-ethylhexyl)phthalate	21.61	149	1944	0.32	ng	# 76
30) Indeno(1,2,3-cd)pyrene	26.56	276	25879	0.32	ng	99
32) Benzo(b)fluoranthene	23.35	252	5803	0.34	ng	99
33) Benzo(k)fluoranthene	23.39	252	7187	0.38	ng	96
34) Benzo(a)pyrene	23.97	252	5660	0.34	ng	97
35) Dibenzo(a,h)anthracene	26.60	278	5050	0.32	ng	96
36) Benzo(g,h,i)perylene	27.33	276	23250	0.35	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE010617\
Data File : BE092630.D
Acq On : 6 Jan 2017 13:47
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM

Quant Time: Jan 07 03:31:01 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 19 16:16:32 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. -0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
152	100		
150	172.4	106.0	159.0#
115	78.9	31.2	46.8#

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM



#2

1,4-Dioxane

Concen: 0.22 ng m




RT: 3.39 min Scan# 69


Delta R.T. 0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
88	100		
58	101.3	63.6	95.4#
43	43.5	28.0	42.0#





#3





n-Nitrosodimethylamine
Concen: 0.40 ng
RT: 3.80 min Scan# 105
Delta R.T. 0.01 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
42	100		
74	115.7	86.0	129.0
44	18.9	5.1	7.7#

Manual Integrations
APPROVED




Sohil
1/9/2017 10:46:21 AM




#4

2-Fluorophenol
Concen: 0.40 ng
RT: 5.69 min Scan# 367
Delta R.T. 0.01 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
112	100		
64	76.4	53.5	80.3
63	30.8	27.9	41.9





#5

Phenol-d6

Concen: 0.42 ng

RT: 7.38 min Scan# 542

Delta R.T. 0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4




Tot Ion:	99	Resp:	662
Ion	Ratio	Lower	Upper
99	100		
42	0.0	16.0	24.0#
71	0.0	30.6	45.8#

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 0.42 ng m

RT: 7.59 min Scan# 551




Delta R.T. -0.00 min


Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Tgt Ion:	93	Resp:	810
Ion	Ratio	Lower	Upper
93	100		
63	52.7	71.6	107.4#
95	0.0	24.0	36.0#

Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.93 min Scan# 707

Delta R.T. 0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
136	100		
137	11.6	8.2	12.4
54	7.1	9.6	14.4#
68	5.1	6.7	10.1#

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM



#8

Nitrobenzene-d5

Concen: 0.43 ng m




RT: 9.36 min Scan# 628


Delta R.T. 0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
82	100		
128	48.9	39.0	58.4
54	73.9	44.8	67.2#





#9

Naphthalene

Concen: 0.38 ng

RT: 10.97 min Scan# 709

Delta R.T. 0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	128	Resp:	3314
Ion	Ratio	Lower	Upper
128	100		
129	14.3	11.2	16.8
127	16.7	11.6	17.4

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM



#10

Hexachlorobutadiene

Concen: 0.38 ng




RT: 11.26 min Scan# 724


Delta R.T. 0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Tgt Ion:	225	Resp:	505
Ion	Ratio	Lower	Upper
225	100		
223	62.2	50.6	76.0
227	62.6	51.4	77.0









#11
2-Methylnaphthalene
Concen: 0.35 ng
RT: 12.64 min Scan# 815
Delta R.T. 0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion:142 Resp: 1952
Ion Ratio Lower Upper
142 100
141 89.4 73.7 110.5
115 42.0 34.0 51.0




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.80 min Scan# 978
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion:164 Resp: 29187
Ion Ratio Lower Upper
164 100
162 94.5 71.4 107.0
160 39.6 27.4 41.2









#13
 2,4,6-Tribromophenol
 Concen: 0.42 ng
 RT: 16.31 min Scan# 1087
 Delta R.T. 0.00 min
 Lab File: BE092630.D
 Acq: 6 Jan 2017 13:47

Instrument :
 BNA_E
 ClientSampled :
 SSTDCCC0.4

Tot Ion:330 Resp: 209
 Ion Ratio Lower Upper
 330 100
 332 90.9 75.4 113.0
 141 36.4 31.0 46.6




Manual Integrations
 APPROVED


Sohil
 1/9/2017 10:46:21 AM



#14
 2-Fluorobiphenyl
 Concen: 0.37 ng
 RT: 13.43 min Scan# 884
 Delta R.T. 0.00 min
 Lab File: BE092630.D
 Acq: 6 Jan 2017 13:47

Tgt Ion:172 Resp: 2581
 Ion Ratio Lower Upper
 172 100
 171 36.5 30.1 45.1
 170 25.3 21.8 32.6









#15
Acenaphthylene
Concen: 0.36 ng
RT: 14.51 min Scan# 961
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
152	100		
151	5.0	3.9	5.9
153	13.0	10.4	15.6




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#16
Acenaphthene
Concen: 0.37 ng
RT: 14.85 min Scan# 981
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
154	100		
153	436.7	350.8	526.2
152	229.7	171.1	256.7





#17

Fluorene

Concen: 0.37 ng

RT: 15.86 min Scan# 1048

Delta R.T. -0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	166	Resp:	2996
Ion Ratio	Lower	Upper	
166	100		
165	98.3	78.2	117.4
167	13.4	11.0	16.4

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.56 min Scan# 1152


Delta R.T. -0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Tgt Ion:	188	Resp:	63859
Ion Ratio	Lower	Upper	
188	100		
94	4.3	3.2	4.8
80	3.9	2.6	3.8#









#19
Hexachlorobenzene
Concen: 0.36 ng/ml
RT: 16.87 min Scan# 1122
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	43.1	27.9	41.9#




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#20
Pentachlorophenol
Concen: 0.41 ng/ml
RT: 17.23 min Scan# 1138
Delta R.T. 0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
266	100		
268	80.3	62.5	93.7
264	82.9	57.2	85.8









#21
Phenanthrene
Concen: 0.35 ng
RT: 17.58 min Scan# 1153
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.8	23.8
179	15.8	16.0	24.0




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#22
Anthracene
Concen: 0.34 ng
RT: 17.69 min Scan# 1158
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.4
179	15.9	12.2	18.2





#23

Fluoranthene

Concen: 0.35 ng

RT: 19.59 min Scan# 1264

Delta R.T. -0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
202	100		
101	2.8	2.2	3.4
203	17.3	13.7	20.5

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.72 min Scan# 1373


Delta R.T. -0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
240	100		
120	2.6	2.6	4.0#
236	26.3	24.6	37.0









#25
Pvrene
Concen: 0.34 ng
RT: 19.95 min Scan# 1285
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
202	100		
200	5.3	4.2	6.4
203	17.4	13.9	20.9




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#26
Terphenyl-d14
Concen: 0.35 ng
RT: 20.17 min Scan# 1298
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.2	6.7	10.1#
122	7.9	3.6	5.4#









#27
Benzo(a)anthracene
Concen: 0.35 ng m
RT: 21.70 min Scan# 1372
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
228	100		
226	26.1	23.6	35.4
229	19.0	13.3	19.9




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#28
Chrysene
Concen: 0.36 ng m
RT: 21.75 min Scan# 1374
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
228	100		
226	37.0	36.3	54.5
229	15.0	10.6	16.0





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.32 ng

RT: 21.61 min Scan# 1368

Delta R.T. 0.00 min

Lab File: BE092630.D

Acq: 6 Jan 2017 13:47

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
149	100		
167	22.2	16.0	24.0
279	0.0	16.0	24.0

Manual Integrations
APPROVED




Sohil
1/9/2017 10:46:21 AM



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.32 ng

RT: 26.56 min Scan# 1713

Delta R.T. -0.00 min

Lab File: BE092630.D




Acq: 6 Jan 2017 13:47


Tgt Ion	Ratio	Lower	Upper
276	100		
138	26.4	20.3	30.5
277	24.7	19.7	29.5

Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.08 min Scan# 1542
Delta R.T. 0.01 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
264	100		
260	25.0	20.1	30.1
265	22.3	17.9	26.9




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#32
Benzo(b)fluoranthene
Concen: 0.34 ng
RT: 23.35 min Scan# 1479
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.2	18.7	28.1
125	12.8	10.5	15.7









#33
Benzo(k)fluoranthene
Concen: 0.38 ng
RT: 23.39 min Scan# 1483
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.1	20.3	30.5
125	12.9	9.6	14.4




Manual Integrations
APPROVED


Sohil
1/9/2017 10:46:21 AM



#34
Benzo(a)pyrene
Concen: 0.34 ng
RT: 23.97 min Scan# 1533
Delta R.T. 0.01 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.8	19.0	28.6
125	13.9	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 0.32 ng
RT: 26.60 min Scan# 1715
Delta R.T. 0.02 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Instrument :
BNA_E
ClientSampled :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
278	100		
139	20.5	13.8	20.8
279	27.7	21.4	32.2

Manual Integrations
APPROVED

Sohil
1/9/2017 10:46:21 AM



#36
Benzo(g,h,i)perylene
Concen: 0.35 ng
RT: 27.33 min Scan# 1758
Delta R.T. -0.00 min
Lab File: BE092630.D
Acq: 6 Jan 2017 13:47

Tgt Ion	Ratio	Lower	Upper
276	100		
277	26.3	19.8	29.8
138	26.2	19.8	29.6

