

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Chemtech Consulting Group Contract: AREC01
 Lab Code: CHEM Case No.: 44964 SAS No.: F5253 SDG NO.: F5253-01
 EPA Sample No.: SSTD0.459 Date Analyzed: 1/7/2015
 Lab File ID: BE089326.D Time Analyzed: 11:05
 Instrument ID: BNA_E GC Column: RXI-5 ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	1259	6.95	0	0.00	0	0.00
UPPER LIMIT	2518	7.45	0	0.00	0	0.00
LOWER LIMIT	630	6.45	0	0.00	0	0.00
EPA SAMPLE NO.						
01 PB81219BL	1223	6.95	0	0.00	0	0.00
02 F5245-15	1219	6.95	0	0.00	0	0.00
03 F5245-16	1207	6.95	0	0.00	0	0.00
04 F5245-17	1259	6.95	0	0.00	0	0.00
05 F5245-18	1289	6.95	0	0.00	0	0.00
06 F5245-19	1269	6.95	0	0.00	0	0.00
07 F5245-20	1158	6.95	0	0.00	0	0.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Chemtech Consulting Group Contract: AREC01
 Lab Code: CHEM Case No.: 44964 SAS No.: F5253 SDG NO.: F5253-01
 EPA Sample No.: SSTD0.460 Date Analyzed: 1/7/2015
 Lab File ID: BE089334.D Time Analyzed: 15:43
 Instrument ID: BNA_E GC Column: RXI-5 ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	1303	6.95	0	0.00	0	0.00
UPPER LIMIT	2606	7.45	0	0.00	0	0.00
LOWER LIMIT	652	6.45	0	0.00	0	0.00
EPA SAMPLE NO.						

IS1 (DCB) = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.