

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010715\
 Data File : BE089344.D
 Acq On : 8 Jan 2015 00:19
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE010715

Manual Integrations
 APPROVED

TEJASKUMAR
 1/9/2015 5:21:13 PM

Quant Time: Jan 08 06:11:01 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BASE-BE010715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 08 06:03:23 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	15048	5.00	ng	0.00
2) Naphthalene-d8	8.51	136	57832	5.00	ng	0.00
6) Acenaphthene-d10	10.65	164	26898	5.00	ng	0.00
11) Phenanthrene-d10	12.72	188	40118	5.00	ng	0.00
15) Chrysene-d12	18.06	240	30491	5.00	ng	0.00
21) Perylene-d12	21.12	264	27041	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	7.65	82	4227	1.05	ng	0.00
7) 2-Fluorobiphenyl	9.85	172	7623	1.04	ng	0.00
17) Terphenyl-d14	15.85	244	5441	1.01	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	8.53	128	13714	1.05	ng	100
5) 2-Methylnaphthalene	9.39	142	8545	1.05	ng	94
8) Acenaphthylene	10.47	152	48068	1.04	ng	99
9) Acenaphthene	10.69	154	7305	1.02	ng	96
10) Fluorene	11.34	166	8040	1.05	ng	99
12) Phenanthrene	12.75	178	43776	1.05	ng	99
13) Anthracene	12.84	178	42675	1.06	ng	99
14) Fluoranthene	14.96	202	9727	1.05	ng	99
16) Pyrene	15.40	202	10327	1.03	ng	99
18) Benzo(a)anthracene	18.05	228	31697	1.03	ng	99
19) Chrysene	18.12	228	32706	1.05	ng	100
20) Indeno(1,2,3-cd)pyrene	23.22	276	30296m	1.01	ng	
22) Benzo(b)fluoranthene	20.36	252	6228	1.05	ng	99
23) Benzo(k)fluoranthene	20.42	252	8148	1.09	ng	100
24) Benzo(a)pyrene	21.00	252	6024	1.03	ng	100
25) Dibenzo(a,h)anthracene	23.30	278	5544	1.04	ng	98
26) Benzo(g,h,i)perylene	23.80	276	27954	1.04	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010715\
Data File : BE089344.D
Acq On : 8 Jan 2015 00:19
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleID :
ICVBE010715

Manual Integrations
APPROVED

TEJASKUMAR
1/9/2015 5:21:13 PM

Quant Time: Jan 08 06:11:01 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BASE-BE010715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 08 06:03:23 2015
Response via : Initial Calibration

