

Data Path : Z:\HPCHEM1\BNA_E\Data\BE010715\
 Data File : BE089334.D
 Acq On : 7 Jan 2015 15:43
 Operator : TP/IZ
 Sample : SSTD0.460
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.460

Quant Time: Jan 08 02:46:19 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 08 01:37:38 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
2 SURR	1,4-Dioxane-d8	0.601	0.604	-0.5	61	-0.02
3	1,4-Dioxane	0.670	0.670	0.0	61	-0.03
4 I	Naphthalene-d8	1.000	1.000	0.0	65	0.00
5	Naphthalene	0.950	0.985	-3.7	62	0.00
6 SURR	2-Methylnaphthalene-d10	0.459	0.485	-5.7	63	0.00
7	2-Methylnaphthalene	0.580	0.629	-8.4	65	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	67	0.00
9	Acenaphthylene	6.314	6.952	-10.1	68	0.00
10 C	Acenaphthene	4.328	4.390	-1.4	64	0.00
11	Fluorene	1.146	1.200	-4.7	65	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
13	Pentachlorophenol	0.041	0.002	95.1#	2#	0.11
14	Phenanthrene	4.281	4.366	-2.0	65	0.00
15	Anthracene	3.824	4.251	-11.2	71	0.00
16 SURR	Fluoranthene-d10	0.693	0.738	-6.5	67	0.00
17 C	Fluoranthene	3.812	3.914	-2.7	64	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	62	0.00
19	Pyrene	5.160	5.842	-13.2	65	0.00
20	Benzo(a)anthracene	0.902	1.058	-17.3	67	0.00
21	Chrysene	1.093	1.078	1.4	57	0.00
22 I	Perylene-d12	1.000	1.000	0.0	66	0.01
23	Benzo(b)fluoranthene	1.026	0.979	4.6	57	0.01
24	Benzo(k)fluoranthene	1.319	1.252	5.1	59	0.01
25 C	Benzo(a)pyrene	1.026	1.103	-7.5	64	0.01
26	Indeno(1,2,3-cd)pyrene	2.822	3.079	-9.1	66	0.01
27	Dibenzo(a,h)anthracene	0.832	0.899	-8.1	67	0.00
28	Benzo(g,h,i)perylene	4.060	4.408	-8.6	67	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0