

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010715\
 Data File : BE089329.D
 Acq On : 7 Jan 2015 12:56
 Operator : TP/IZ
 Sample : F5245-16
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 Y9RA0

Quant Time: Jan 08 13:03:08 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 02 10:52:56 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	1207	0.40	ng/ul	0.00
4) Naphthalene-d8	8.51	136	4739	0.40	ng/ul	0.00
8) Acenaphthene-d10	10.66	164	2349	0.40	ng/ul	0.00
12) Phenanthrene-d10	12.71	188	3706	0.40	ng/ul	0.00
18) Chrysene-d12	18.07	240	2722	0.40	ng/ul	0.00
22) Perylene-d12	21.12	264	2452	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.12	96	442	0.24	ng/uL	-0.03
6) 2-Methylnaphthalene-d10	0.00	152	0	0.00	ng/ul	
16) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
						Qvalue
17) Fluoranthene	14.96	202	878	0.02	ng/ul#	59
19) Pyrene	15.41	202	941	0.03	ng/ul#	60


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010715\
Data File : BE089329.D
Acq On : 7 Jan 2015 12:56
Operator : TP/IZ
Sample : F5245-16
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
Y9RA0

Quant Time: Jan 08 13:03:08 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 02 10:52:56 2015
Response via : Initial Calibration









#17
Fluoranthene
Concen: 0.02 ng/ul
RT: 14.96 min Scan# 2030
Delta R.T. 0.01 min
Lab File: BE089329.D
Acq: 7 Jan 2015 12:56

Instrument :
BNA_E
ClientSampleId :
Y9RA0

Tgt Ion:202 Resp: 878
Ion Ratio Lower Upper
202 100
101 10.4 0.0 24.6
203 38.5 0.0 37.8#



#19
Pyrene
Concen: 0.03 ng/ul
RT: 15.41 min Scan# 2099
Delta R.T. 0.01 min
Lab File: BE089329.D
Acq: 7 Jan 2015 12:56

Tgt Ion:202 Resp: 941
Ion Ratio Lower Upper
202 100
200 9.9 4.2 6.4#
203 39.1 14.4 21.6#

