

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010815\
 Data File : BE089370.D
 Acq On : 8 Jan 2015 17:40
 Operator : TP/IZ
 Sample : F5245-15RX
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 Y9R92

Quant Time: Jan 09 12:06:52 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 09 01:27:36 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	1237	0.40	ng/ul	0.00
4) Naphthalene-d8	8.51	136	4976	0.40	ng/ul	0.00
8) Acenaphthene-d10	10.65	164	2533	0.40	ng/ul	0.00
12) Phenanthrene-d10	12.71	188	3830	0.40	ng/ul	0.00
18) Chrysene-d12	18.07	240	2652	0.40	ng/ul	0.00
22) Perylene-d12	21.12	264	2271	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.12	96	503	0.27	ng/uL	0.01
6) 2-Methylnaphthalene-d10	0.00	152	0	0.00	ng/ul	
16) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
						Qvalue
17) Fluoranthene	14.96	202	983	0.03	ng/ul#	59
19) Pyrene	15.40	202	1077	0.03	ng/ul#	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010815\
Data File : BE089370.D
Acq On : 8 Jan 2015 17:40
Operator : TP/IZ
Sample : F5245-15RX
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
Y9R92

Quant Time: Jan 09 12:06:52 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 09 01:27:36 2015
Response via : Initial Calibration

