

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010820\
 Data File : BE101008.D
 Acq On : 8 Jan 2020 18:41
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE010820

Quant Time: Jan 09 10:52:04 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	-0.01
2	1,4-Dioxane	1.020	0.916	10.2	101	0.00
3	n-Nitrosodimethylamine	0.556	0.586	-5.4	127	-0.01
4 S	2-Fluorophenol	0.900	0.885	1.7	103	0.00
5 S	Phenol-d6	1.139	1.112	2.4	114	-0.01
6	bis(2-Chloroethyl)ether	1.357	1.241	8.5	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.286	0.268	6.3	110	0.00
9	Naphthalene	4.441	4.563	-2.7	106	0.00
10	Hexachlorobutadiene	0.188	0.197	-4.8	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.763	0.780	-2.2	107	0.00
12	2-Methylnaphthalene	0.672	0.679	-1.0	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.113	0.104	8.0	112	0.00
15 S	2-Fluorobiphenyl	1.519	1.562	-2.8	106	-0.01
16	Acenaphthylene	1.629	1.506	7.6	107	0.00
17	Acenaphthene	1.166	1.184	-1.5	109	0.00
18	Fluorene	1.474	1.504	-2.0	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.195	0.190	2.6	112	0.00
21	Hexachlorobenzene	0.214	0.212	0.9	107	0.00
22	Pentachlorophenol	0.053	0.042	20.8	114	0.00
23	Phenanthrene	1.092	1.092	0.0	111	0.00
24	Anthracene	1.031	0.991	3.9	111	0.00
25 SURR	Fluoranthene-d10	5.555	5.209	6.2	106	0.00
26	Fluoranthene	1.368	1.292	5.6	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pvrene	1.489	1.577	-5.9	106	0.00
29 S	Terphenyl-d14	0.982	1.034	-5.3	107	0.00
30	Benzo(a)anthracene	1.141	0.991	13.1	93	0.00
31	Chrysene	1.696	1.850	-9.1	109	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	1.702	10.5	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.308	2.4	103	0.00
34 I	Perylene-d12	1.000	1.000	0.0	106	0.00
35	Benzo(b)fluoranthene	1.122	1.044	7.0	105	0.00
36	Benzo(k)fluoranthene	1.626	1.493	8.2	97	0.00
37 C	Benzo(a)pyrene	1.085	1.016	6.4	102	0.00
38	Dibenzo(a,h)anthracene	0.998	0.943	5.5	108	0.00
39	Benzo(a,h,i)perylene	1.227	1.165	5.1	98	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010820\
Data File : BE101008.D
Acq On : 8 Jan 2020 18:41
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE010820

Quant Time: Jan 09 10:52:04 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			