

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010820\
 Data File : BE101001.D
 Acq On : 8 Jan 2020 13:24
 Operator : JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 1/9/2020 12:12:47 PM

Quant Time: Jan 08 18:28:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:14:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	3373	0.40	ng	0.00
7) Naphthalene-d8	10.51	136	15512	0.40	ng	0.01
13) Acenaphthene-d10	14.37	164	9440	0.40	ng	0.01
19) Phenanthrene-d10	17.09	188	21652	0.40	ng	0.00
27) Chrysene-d12	21.28	240	17384	0.40	ng	0.00
34) Perylene-d12	23.70	264	20364	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.34	112	1001	0.18	ng	0.00
5) Phenol-d6	6.92	99	999	0.12	ng	0.00
8) Nitrobenzene-d5	8.87	82	1019	0.14	ng	0.00
11) 2-Methylnaphthalene-d10	12.11	152	3089	0.14	ng	0.00
14) 2,4,6-Tribromophenol	15.85	330	233	0.11	ng	0.00
15) 2-Fluorobiphenyl	13.00	172	3589	0.11	ng	0.00
25) Fluoranthene-d10	19.13	212	27760	0.11	ng	0.00
29) Terphenyl-d14	19.74	244	4274	0.11	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.30	88	1473	0.124	ng	# 82
6) bis(2-Chloroethyl)ether	7.16	93	1227	0.101	ng	98
9) Naphthalene	10.55	128	16860	0.102	ng	# 98
10) Hexachlorobutadiene	10.85	225	783	0.107	ng	99
12) 2-Methylnaphthalene	12.18	142	2481	0.099	ng	98
16) Acenaphthylene	14.08	152	3402	0.096	ng	97
17) Acenaphthene	14.43	154	2692	0.105	ng	100
18) Fluorene	15.41	166	3366	0.105	ng	99
20) 4-Bromophenyl-phenylether	16.31	248	1018	0.107	ng	98
21) Hexachlorobenzene	16.41	284	1126	0.100	ng	97
23) Phenanthrene	17.14	178	5752	0.103	ng	98
24) Anthracene	17.23	178	5092	0.090	ng	99
26) Fluoranthene	19.16	202	6692	0.088	ng	98
28) Pyrene	19.52	202	6793	0.112	ng	100
30) Benzo(a)anthracene	21.27	228	4578m	0.108	ng	
31) Chrysene	21.32	228	8070	0.102	ng	97
32) Bis(2-ethylhexyl)phthalate	21.20	149	8490	0.091	ng	# 99
33) Indeno(1,2,3-cd)pyrene	26.25	276	5965	0.100	ng	# 82
35) Benzo(b)fluoranthene	22.96	252	4695	0.110	ng	# 89
36) Benzo(k)fluoranthene	23.01	252	8043m	0.099	ng	
37) Benzo(a)pyrene	23.60	252	4957	0.099	ng	# 80
38) Dibenzo(a,h)anthracene	26.28	278	3902	0.086	ng	# 86
39) Benzo(g,h,i)perylene	27.03	276	5462	0.102	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010820\
Data File : BE101001.D
Acq On : 8 Jan 2020 13:24
Operator : JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleID :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM

Quant Time: Jan 08 18:28:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:14:32 2020
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.73 min Scan# 423
Delta R.T. -0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
152	100		
150	153.0	120.3	180.5
115	67.7	51.7	77.5

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane
Concen: 0.124 ng
RT: 3.30 min Scan# 38
Delta R.T. 0.01 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
88	100		
58	45.7	53.0	79.4#
43	29.0	23.4	35.2






Abundance

Ion 88.00 (87.70 to 88.70): BE1

Ion 58.00 (57.70 to 58.70): BE1

Ion 43.00 (42.70 to 43.70): BE1





#4

2-Fluorophenol

Concen: 0.176 ng

RT: 5.34 min Scan# 215

Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tot Ion:	112	Resp:	1001
Ion	Ratio	Lower	Upper
112	100		
64	69.1	53.6	80.4
63	129.5	114.6	172.0

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM



#5

Phenol-d6

Concen: 0.125 ng




RT: 6.92 min Scan# 352


Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Tgt Ion:	99	Resp:	999
Ion	Ratio	Lower	Upper
99	100		
42	25.6	18.3	27.5
71	0.0	28.0	42.0#





#6

bis(2-Chloroethyl)ether

Concen: 0.101 ng

RT: 7.16 min Scan# 373

Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tot Ion:	93	Resp:	1227
Ion	Ratio	Lower	Upper
93	100		
63	288.8	233.9	350.9
95	31.0	23.5	35.3

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM



#7

Naphthalene-d8

Concen: 0.400 ng




RT: 10.51 min Scan# 643


Delta R.T. 0.01 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Tgt Ion:	136	Resp:	15512
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.3	13.9
54	30.3	36.8	55.2#
68	7.0	8.4	12.6#





#8

Nitrobenzene-d5

Concen: 0.137 ng

RT: 8.87 min Scan# 517

Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
82	100		
128	171.1	109.8	164.6#
54	276.5	174.5	261.7#

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM






Abundance

Ion 82.00 (81.70 to 82.70): BE1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE1



#9

Naphthalene

Concen: 0.102 ng


RT: 10.55 min Scan# 646

Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
128	100		
129	3.6	2.3	3.5#
127	4.3	2.7	4.1#






Abundance

Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E






#10
Hexachlorobutadiene
Concen: 0.107 ng
RT: 10.85 min Scan# 669
Delta R.T. -0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
225	100		
223	64.1	52.2	78.4
227	62.1	49.9	74.9

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM






Abundance

Ion 225.00 (224.70 to 225.70): E


Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#11
2-Methylnaphthalene-d10
Concen: 0.136 ng
RT: 12.11 min Scan# 757
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24



Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.0	22.6




Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E





#12

2-Methylnaphthalene

Concen: 0.099 ng

RT: 12.18 min Scan# 762

Delta R.T. 0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E


ClientSampleID :

SSTDIC0.1

Tot Ion:	142	Resp:	2481
Ion	Ratio	Lower	Upper
142	100		
141	89.5	72.5	108.7
115	44.0	33.0	49.6

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#13

Acenaphthene-d10

Concen: 0.400 ng


RT: 14.37 min Scan# 914

Delta R.T. 0.01 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Tgt Ion:	164	Resp:	9440
Ion	Ratio	Lower	Upper
164	100		
162	103.9	84.1	126.1
160	46.7	40.4	60.6






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#14

2,4,6-Tribromophenol

Concen: 0.112 ng

RT: 15.85 min Scan# 1018

Delta R.T. 0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
330	100		
332	114.6	78.4	117.6
141	47.2	37.0	55.4

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM



#15

2-Fluorobiphenyl

Concen: 0.107 ng




RT: 13.00 min Scan# 819


Delta R.T. 0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.1	28.2	42.2
170	28.4	18.9	28.3#









#16
Acenaphthylene
Concen: 0.096 ng
RT: 14.08 min Scan# 894
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
152	100		
151	21.1	15.5	23.3
153	12.3	10.4	15.6




Manual Integrations
APPROVED


mohammad
1/9/2020 12:12:47 PM



#17
Acenaphthene
Concen: 0.105 ng
RT: 14.43 min Scan# 918
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
154	100		
153	116.6	93.6	140.4
152	57.3	46.3	69.5









#18
Fluorene
 Concen: 0.105 ng
 RT: 15.41 min Scan# 985
 Delta R.T. 0.00 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
166	100		
165	99.4	79.8	119.8
167	16.1	11.0	16.6




Manual Integrations
 APPROVED


mohammad
 1/9/2020 12:12:47 PM



#19
Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.09 min Scan# 1111
 Delta R.T. 0.00 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	13.9	10.8	16.2






#20
4-Bromophenyl-phenylether
Concen: 0.107 ng
RT: 16.31 min Scan# 1052
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
248	100		
250	95.2	76.6	115.0
141	75.4	57.7	86.5

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM




Abundance



Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



Time-->



#21
Hexachlorobenzene
Concen: 0.100 ng
RT: 16.41 min Scan# 1060
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24


Tgt Ion	Ratio	Lower	Upper
284	100		
142	41.4	34.0	51.0
249	32.5	28.2	42.4

Abundance



Ion 283.80 (283.50 to 284.50): E


Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



Time-->





#23

Phenanthrene

Concen: 0.103 ng

RT: 17.14 min Scan# 1114

Delta R.T. 0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E


ClientSampleId :

SSTDICC0.1




Tot Ion:178	Resp:	5752
Ion Ratio	Lower	Upper
178	100	
176	20.5	15.7 23.5
179	16.2	12.0 18.0

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E



#24

Anthracene

Concen: 0.090 ng

RT: 17.23 min Scan# 1121




Delta R.T. 0.00 min


Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Tgt Ion:178	Resp:	5092
Ion Ratio	Lower	Upper
178	100	
176	18.5	14.8 22.2
179	14.4	12.4 18.6

Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E









#25
 Fluoranthene-d10
 Concen: 0.106 ng
 RT: 19.13 min Scan# 1370
 Delta R.T. -0.00 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
212	100		
106	3.2	2.6	3.8
104	1.9	1.4	2.0




Manual Integrations
 APPROVED


mohammad
 1/9/2020 12:12:47 PM



#26
 Fluoranthene
 Concen: 0.088 ng
 RT: 19.16 min Scan# 1375
 Delta R.T. -0.00 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
202	100		
101	11.2	9.6	14.4
203	16.1	13.8	20.6





#27

Chrysene-d12

Concen: 0.400 ng

RT: 21.28 min Scan# 1624

Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
240	100		
120	10.5	8.6	13.0
236	26.8	22.2	33.4

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM






Abundance

Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#28

Pyrene

Concen: 0.112 ng


RT: 19.52 min Scan# 1437

Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.4	17.0	25.4
203	17.4	13.8	20.6






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#29
 Terphenyl-d14
 Concen: 0.107 ng
 RT: 19.74 min Scan# 1475
 Delta R.T. -0.00 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
244	100		
212	38.0	27.2	40.8
122	14.4	9.8	14.8




Manual Integrations
 APPROVED


mohammad
 1/9/2020 12:12:47 PM



#30
 Benzo(a)anthracene
 Concen: 0.108 ng m
 RT: 21.27 min Scan# 1623
 Delta R.T. 0.01 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.0	22.7	34.1
229	23.7	15.5	23.3#









#31
 Chrysene
 Concen: 0.102 ng
 RT: 21.32 min Scan# 1627
 Delta R.T. -0.00 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
228	100		
226	30.9	23.5	35.3
229	22.1	16.7	25.1




Manual Integrations
 APPROVED


mohammad
 1/9/2020 12:12:47 PM



#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.091 ng
 RT: 21.20 min Scan# 1617
 Delta R.T. -0.00 min
 Lab File: BE101001.D
 Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
149	100		
167	6.8	5.2	7.8
279	1.2	0.8	1.2#






#33
Indeno(1,2,3-cd)pyrene
Concen: 0.100 ng
RT: 26.25 min Scan# 2864
Delta R.T. -0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tot Ion	Ratio	Lower	Upper
276	100		
138	17.4	16.1	24.1
277	9.0	18.1	27.1#

Manual Integrations
APPROVED



mohammad
1/9/2020 12:12:47 PM



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E




Time-->





#34
Perylene-d12
Concen: 0.400 ng
RT: 23.70 min Scan# 2154
Delta R.T. -0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24


Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.0	20.3	30.5
265	32.1	27.0	40.4

Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E



Time-->









#35
Benzo(b)fluoranthene
Concen: 0.110 ng
RT: 22.96 min Scan# 1947
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
252	100		
253	27.3	18.5	27.7
125	18.4	9.5	14.3#




Manual Integrations
APPROVED


mohammad
1/9/2020 12:12:47 PM



#36
Benzo(k)fluoranthene
Concen: 0.099 ng m
RT: 23.01 min Scan# 1960
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
252	100		
253	28.0	18.9	28.3
125	22.3	10.1	15.1#









#37
Benzo(a)pyrene
Concen: 0.099 ng
RT: 23.60 min Scan# 2125
Delta R.T. 0.01 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
252	100		
253	30.5	19.4	29.2#
125	27.9	11.8	17.6#




Manual Integrations
APPROVED


mohammad
1/9/2020 12:12:47 PM



#38
Dibenzo(a,h)anthracene
Concen: 0.086 ng
RT: 26.28 min Scan# 2873
Delta R.T. 0.00 min
Lab File: BE101001.D
Acq: 8 Jan 2020 13:24

Tgt Ion	Ratio	Lower	Upper
278	100		
139	28.9	15.0	22.4#
279	32.6	22.6	33.8





#39

Benzo(a,h,i)perylene

Concen: 0.102 ng

RT: 27.03 min Scan# 3082

Delta R.T. -0.00 min

Lab File: BE101001.D

Acq: 8 Jan 2020 13:24

Instrument :

BNA_E

ClientSampleId :

SSTDICC0.1

Tot Ion:	276	Resp:	5462
Ion Ratio	Lower	Upper	
276	100		
277	27.6	19.4	29.2
138	28.6	17.9	26.9#

Manual Integrations
APPROVED

mohammad
1/9/2020 12:12:47 PM

