

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010820\
 Data File : BE101007.D
 Acq On : 8 Jan 2020 17:08
 Operator : JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Manual Integrations
 APPROVED

mohammad
 1/9/2020 12:13:02 PM

Quant Time: Jan 08 18:42:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:14:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	4728	0.40	ng	0.00
7) Naphthalene-d8	10.50	136	18903	0.40	ng	0.00
13) Acenaphthene-d10	14.36	164	11322	0.40	ng	0.00
19) Phenanthrene-d10	17.09	188	24240	0.40	ng	0.00
27) Chrysene-d12	21.27	240	24746	0.40	ng	-0.01
34) Perylene-d12	23.70	264	23051	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.33	112	45732	5.72	ng	0.00
5) Phenol-d6	6.90	99	72117	6.43	ng	-0.02
8) Nitrobenzene-d5	8.86	82	76468	8.44	ng	-0.01
11) 2-Methylnaphthalene-d10	12.09	152	167279	6.02	ng	-0.01
14) 2,4,6-Tribromophenol	15.84	330	19312	7.77	ng	-0.01
15) 2-Fluorobiphenyl	12.99	172	200005	4.97	ng	-0.01
25) Fluoranthene-d10	19.13	212	1704589	5.79	ng	0.00
29) Terphenyl-d14	19.73	244	283135	4.98	ng	0.00

Target Compounds				Qvalue		
3) n-Nitrosodimethylamine	3.59	42	30450	7.803	ng	# 90
6) bis(2-Chloroethyl)ether	7.15	93	69677	4.077	ng	88
9) Naphthalene	10.55	128	961930	4.786	ng	100
10) Hexachlorobutadiene	10.85	225	39177	4.411	ng	98
12) 2-Methylnaphthalene	12.17	142	159270	5.230	ng	99
16) Acenaphthylene	14.08	152	255723	6.043	ng	100
17) Acenaphthene	14.41	154	163318	5.312	ng	96
18) Fluorene	15.39	166	208217	5.434	ng	99
20) 4-Bromophenyl-phenylether	16.29	248	60386	5.674	ng	99
21) Hexachlorobenzene	16.40	284	61839	4.920	ng	99
22) Pentachlorophenol	16.75	266	21929	11.681	ng	86
23) Phenanthrene	17.14	178	330084	5.283	ng	99
24) Anthracene	17.23	178	331314	5.244	ng	100
26) Fluoranthene	19.16	202	424955	4.988	ng	99
28) Pyrene	19.52	202	421145	4.865	ng	100
30) Benzo(a)anthracene	21.26	228	368835	6.089	ng	97
31) Chrysene	21.31	228	458575	4.066	ng	97
32) Bis(2-ethylhexyl)phthalate	21.20	149	699369	5.249	ng	100
33) Indeno(1,2,3-cd)pyrene	26.24	276	407341	4.798	ng	# 93
35) Benzo(b)fluoranthene	22.95	252	357821	7.374	ng	97
36) Benzo(k)fluoranthene	23.00	252	447114m	4.860	ng	
37) Benzo(a)pyrene	23.59	252	335157	5.916	ng	# 94
38) Dibenzo(a,h)anthracene	26.26	278	328086	6.391	ng	93
39) Benzo(g,h,i)perylene	27.02	276	354163	5.857	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.73 min Scan# 422
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampled :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
152	100		
150	155.5	120.3	180.5
115	66.1	51.7	77.5

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#3

n-Nitrosodimethylamine
Concen: 7.803 ng
RT: 3.59 min Scan# 63
Delta R.T. -0.02 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
42	100		
74	120.3	105.7	158.5
44	6.9	10.0	15.0#






Abundance

Ion 42.00 (41.70 to 42.70): BE1

Ion 74.00 (73.70 to 74.70): BE1

Ion 44.00 (43.70 to 44.70): BE1





#4

2-Fluorophenol

Concen: 5.721 ng

RT: 5.33 min Scan# 214

Delta R.T. -0.01 min

Lab File: BE101007.D

Acq: 8 Jan 2020 17:08

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion:	112	Resp:	45732
Ion Ratio	Lower	Upper	
112	100		
64	68.3	53.6	80.4
63	143.4	114.6	172.0

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM



#5

Phenol-d6

Concen: 6.426 ng




RT: 6.90 min Scan# 350


Delta R.T. -0.02 min

Lab File: BE101007.D

Acq: 8 Jan 2020 17:08

Tgt Ion:	99	Resp:	72117
Ion Ratio	Lower	Upper	
99	100		
42	21.2	18.3	27.5
71	35.3	28.0	42.0





#6

bis(2-Chloroethyl)ether

Concen: 4.077 ng

RT: 7.15 min Scan# 372

Delta R.T. -0.01 min

Lab File: BE101007.D

Acq: 8 Jan 2020 17:08

Instrument :

BNA_E





ClientSampled :

SSTDICC5.0

Tot Ion:	93	Resp:	69677
Ion	Ratio	Lower	Upper
93	100		
63	316.8	233.9	350.9
95	33.3	23.5	35.3

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM



#7

Naphthalene-d8

Concen: 0.400 ng




RT: 10.50 min Scan# 641


Delta R.T. -0.00 min

Lab File: BE101007.D

Acq: 8 Jan 2020 17:08

Tgt Ion:	136	Resp:	18903
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.3	13.9
54	33.5	36.8	55.2#
68	7.9	8.4	12.6#





#8

Nitrobenzene-d5

Concen: 8.443 ng

RT: 8.86 min Scan# 515

Delta R.T. -0.01 min

Lab File: BE101007.D

Acq: 8 Jan 2020 17:08

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
82	100		
128	118.5	109.8	164.6
54	195.9	174.5	261.7

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM



#9

Naphthalene

Concen: 4.786 ng




RT: 10.55 min Scan# 645


Delta R.T. -0.00 min

Lab File: BE101007.D

Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
128	100		
129	2.8	2.3	3.5
127	3.2	2.7	4.1









#10
Hexachlorobutadiene
Concen: 4.411 ng
RT: 10.85 min Scan# 668
Delta R.T. -0.00 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
225	100		
223	62.9	52.2	78.4
227	63.7	49.9	74.9




Manual Integrations
APPROVED


mohammad
1/9/2020 12:13:02 PM



#11
2-Methylnaphthalene-d10
Concen: 6.024 ng
RT: 12.09 min Scan# 755
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.0	22.6






#12
2-Methylnaphthalene
Concen: 5.230 ng
RT: 12.17 min Scan# 760
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
142	100		
141	90.9	72.5	108.7
115	39.9	33.0	49.6

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.36 min Scan# 912
Delta R.T. 0.00 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08




Tgt Ion	Ratio	Lower	Upper
164	100		
162	105.4	84.1	126.1
160	50.0	40.4	60.6


Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#14

2,4,6-Tribromophenol

Concen: 7.767 ng

RT: 15.84 min Scan# 1016

Delta R.T. -0.01 min

Lab File: BE101007.D

Acq: 8 Jan 2020 17:08

Instrument :

BNA_E


ClientSampled :

SSTDICC5.0




Tot Ion:	330	Resp:	19312
Ion Ratio	Lower	Upper	
330	100		
332	96.5	78.4	117.6
141	48.1	37.0	55.4

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM



Abundance Ion 330.00 (329.70 to 330.70): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E



#15

2-Fluorobiphenyl

Concen: 4.971 ng


RT: 12.99 min Scan# 817

Delta R.T. -0.01 min



Lab File: BE101007.D


Acq: 8 Jan 2020 17:08

Tgt Ion:	172	Resp:	200005
Ion Ratio	Lower	Upper	
172	100		
171	34.6	28.2	42.2
170	23.0	18.9	28.3



Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E









#16
Acenaphthylene
Concen: 6.043 ng
RT: 14.08 min Scan# 893
Delta R.T. 0.00 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.5	23.3
153	13.1	10.4	15.6




Manual Integrations
APPROVED


mohammad
1/9/2020 12:13:02 PM



#17
Acenaphthene
Concen: 5.312 ng
RT: 14.41 min Scan# 916
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
154	100		
153	111.8	93.6	140.4
152	55.4	46.3	69.5









#18
Fluorene
 Concen: 5.434 ng
 RT: 15.39 min Scan# 983
 Delta R.T. -0.01 min
 Lab File: BE101007.D
 Acq: 8 Jan 2020 17:08

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
166	100		
165	99.2	79.8	119.8
167	13.5	11.0	16.6




Manual Integrations
 APPROVED


mohammad
 1/9/2020 12:13:02 PM



#19
Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.09 min Scan# 1110
 Delta R.T. 0.00 min
 Lab File: BE101007.D
 Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.3	10.8	16.2#






#20
4-Bromophenyl-phenylether
Concen: 5.674 ng
RT: 16.29 min Scan# 1050
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
248	100		
250	95.7	76.6	115.0
141	69.4	57.7	86.5

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM






Abundance

Ion 248.00 (247.70 to 248.70): E


Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#21
Hexachlorobenzene
Concen: 4.920 ng
RT: 16.40 min Scan# 1058
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
284	100		
142	41.8	34.0	51.0
249	34.5	28.2	42.4






Abundance

Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E






#22
 Pentachlorophenol
 Concen: 11.681 ng
 RT: 16.75 min Scan# 1084
 Delta R.T. -0.01 min
 Lab File: BE101007.D
 Acq: 8 Jan 2020 17:08

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
266	100		
264	64.4	61.0	91.4
268	61.9	58.5	87.7

Manual Integrations
 APPROVED

mohammad
 1/9/2020 12:13:02 PM






Abundance

Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E



#23
 Phenanthrene
 Concen: 5.283 ng
 RT: 17.14 min Scan# 1113
 Delta R.T. 0.00 min
 Lab File: BE101007.D
 Acq: 8 Jan 2020 17:08




Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.7	23.5
179	15.4	12.0	18.0


Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#24
 Anthracene
 Concen: 5.244 ng
 RT: 17.23 min Scan# 1120
 Delta R.T. 0.00 min
 Lab File: BE101007.D
 Acq: 8 Jan 2020 17:08

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tot Ion	Ion	Ratio	Lower	Upper
178	100			
176	18.7	14.8	22.2	
179	15.3	12.4	18.6	




Manual Integrations
 APPROVED


mohammad
 1/9/2020 12:13:02 PM



#25
 Fluoranthene-d10
 Concen: 5.795 ng
 RT: 19.13 min Scan# 1368
 Delta R.T. -0.01 min
 Lab File: BE101007.D
 Acq: 8 Jan 2020 17:08

Tgt Ion	Ion	Ratio	Lower	Upper
212	100			
106	3.5	2.6	3.8	
104	1.9	1.4	2.0	






#26
Fluoranthene
Concen: 4.988 ng
RT: 19.16 min Scan# 1373
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
202	100		
101	12.5	9.6	14.4
203	17.2	13.8	20.6

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM




Abundance



Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



Time--> 19.00 19.20 19.40



#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.27 min Scan# 1622
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08


Tgt Ion	Ratio	Lower	Upper
240	100		
120	12.2	8.6	13.0
236	27.9	22.2	33.4

Abundance



Ion 240.00 (239.70 to 240.70): E


Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



Time--> 21.20 21.40









#28
Pvrene
Concen: 4.865 ng
RT: 19.52 min Scan# 1436
Delta R.T. -0.00 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.0	25.4
203	17.5	13.8	20.6




Manual Integrations
APPROVED


mohammad
1/9/2020 12:13:02 PM



#29
Terphenyl-d14
Concen: 4.978 ng
RT: 19.73 min Scan# 1473
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
244	100		
212	31.1	27.2	40.8
122	11.9	9.8	14.8









#30
Benzo(a)anthracene
Concen: 6.089 ng
RT: 21.26 min Scan# 1621
Delta R.T. -0.00 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
228	100		
226	26.6	22.7	34.1
229	20.1	15.5	23.3




Manual Integrations
APPROVED


mohammad
1/9/2020 12:13:02 PM



#31
Chrysene
Concen: 4.066 ng
RT: 21.31 min Scan# 1625
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.9	23.5	35.3
229	19.5	16.7	25.1






#32
Bis(2-ethylhexyl)phthalate
Concen: 5.249 ng
RT: 21.20 min Scan# 1616
Delta R.T. -0.00 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
149	100		
167	6.4	5.2	7.8
279	0.8	0.8	1.2

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#33
Indeno(1,2,3-cd)pyrene
Concen: 4.798 ng
RT: 26.24 min Scan# 2860
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08




Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.8	16.1	24.1#
277	24.9	18.1	27.1


Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.70 min Scan# 2151
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
264	100		
260	22.9	20.3	30.5
265	28.9	27.0	40.4




Manual Integrations
APPROVED


mohammad
1/9/2020 12:13:02 PM



#35
Benzo(b)fluoranthene
Concen: 7.374 ng
RT: 22.95 min Scan# 1943
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	18.5	27.7
125	10.6	9.5	14.3









#36
Benzo(k)fluoranthene
Concen: 4.860 ng m
RT: 23.00 min Scan# 1956
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.2	18.9	28.3
125	10.6	10.1	15.1




Manual Integrations
APPROVED


mohammad
1/9/2020 12:13:02 PM



#37
Benzo(a)pyrene
Concen: 5.916 ng
RT: 23.59 min Scan# 2120
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.8	19.4	29.2
125	11.6	11.8	17.6#









#38
Dibenzo(a,h)anthracene
Concen: 6.391 ng
RT: 26.26 min Scan# 2867
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Instrument :
BNA_E
ClientSampled :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
278	100		
139	16.3	15.0	22.4
279	23.6	22.6	33.8

Manual Integrations
APPROVED

mohammad
1/9/2020 12:13:02 PM



#39
Benzo(g,h,i)perylene
Concen: 5.857 ng
RT: 27.02 min Scan# 3079
Delta R.T. -0.01 min
Lab File: BE101007.D
Acq: 8 Jan 2020 17:08

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.4	29.2
138	21.8	17.9	26.9

