

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010915\
 Data File : BE089390.D
 Acq On : 9 Jan 2015 20:39
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Manual Integrations
 APPROVED

tejaskumar
 1/12/2015 4:10:19 PM

Quant Time: Jan 10 00:43:05 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BASE-BE010715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 10 00:23:37 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.93	152	18739	5.00	ng	0.00
2) Naphthalene-d8	8.49	136	72289	5.00	ng	0.00
6) Acenaphthene-d10	10.63	164	33642	5.00	ng	0.00
11) Phenanthrene-d10	12.68	188	51579	5.00	ng	0.00
15) Chrysene-d12	18.02	240	38114	5.00	ng	0.00
21) Perylene-d12	21.06	264	33293	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	7.63	82	5277	1.05	ng	0.00
7) 2-Fluorobiphenyl	9.83	172	9504	1.04	ng	0.00
17) Terphenyl-d14	15.80	244	6598	0.98	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	8.52	128	16776	1.03	ng	99
5) 2-Methylnaphthalene	9.37	142	10476	1.03	ng	99
8) Acenaphthylene	10.45	152	60403	1.05	ng	100
9) Acenaphthene	10.67	154	9027	1.01	ng	93
10) Fluorene	11.32	166	9918	1.03	ng	100
12) Phenanthrene	12.72	178	55813	1.04	ng	100
13) Anthracene	12.80	178	53936	1.04	ng	99
14) Fluoranthene	14.91	202	11984	1.01	ng	99
16) Pyrene	15.36	202	12865	1.03	ng	100
18) Benzo(a)anthracene	18.00	228	37559	0.97	ng	100
19) Chrysene	18.08	228	40723	1.05	ng	99
20) Indeno(1,2,3-cd)pyrene	23.15	276	31826m	0.86	ng	
22) Benzo(b)fluoranthene	20.31	252	6773	0.93	ng	100
23) Benzo(k)fluoranthene	20.37	252	9970	1.09	ng	100
24) Benzo(a)pyrene	20.95	252	6572	0.92	ng	99
25) Dibenzo(a,h)anthracene	23.23	278	5511	0.84	ng	99
26) Benzo(g,h,i)perylene	23.73	276	29026	0.88	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010915\
Data File : BE089390.D
Acq On : 9 Jan 2015 20:39
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Manual Integrations
APPROVED

tejaskumar
1/12/2015 4:10:19 PM

Quant Time: Jan 10 00:43:05 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BASE-BE010715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 10 00:23:37 2015
Response via : Initial Calibration

