

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010915\
 Data File : BE089378.D
 Acq On : 9 Jan 2015 11:03
 Operator : TP/IZ
 Sample : SSTD0.463
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.463

Quant Time: Jan 09 23:23:17 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 09 23:13:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	58	0.00
2 SURR	1,4-Dioxane-d8	0.601	0.600	0.2	54	-0.05
3	1,4-Dioxane	0.670	0.674	-0.6	56	-0.05
4 I	Naphthalene-d8	1.000	1.000	0.0	59	0.00
5	Naphthalene	0.950	0.984	-3.6	57	0.00
6 SURR	2-Methylnaphthalene-d10	0.459	0.481	-4.8	57	0.00
7	2-Methylnaphthalene	0.580	0.635	-9.5	59	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	60	0.00
9	Acenaphthylene	6.314	7.075	-12.1	62	0.00
10 C	Acenaphthene	4.328	4.457	-3.0	59	0.00
11	Fluorene	1.146	1.210	-5.6	59	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	62	0.00
13	Pentachlorophenol	0.041	0.000	100.0#	0#	-12.45#
14	Phenanthrene	4.281	4.406	-2.9	60	0.00
15	Anthracene	3.824	4.147	-8.4	63	0.00
16 SURR	Fluoranthene-d10	0.693	0.716	-3.3	60	0.00
17 C	Fluoranthene	3.812	3.939	-3.3	59	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	58	0.00
19	Pyrene	5.160	5.653	-9.6	59	0.00
20	Benzo(a)anthracene	0.902	0.996	-10.4	59	0.00
21	Chrysene	1.093	1.097	-0.4	54	0.00
22 I	Perylene-d12	1.000	1.000	0.0	64	0.01
23	Benzo(b)fluoranthene	1.026	0.875	14.7	50	0.01
24	Benzo(k)fluoranthene	1.319	1.273	3.5	58	0.01
25 C	Benzo(a)pyrene	1.026	1.060	-3.3	60	0.01
26	Indeno(1,2,3-cd)pyrene	2.822	2.606	7.7	54	0.01
27	Dibenzo(a,h)anthracene	0.832	0.838	-0.7	61	0.00
28	Benzo(g,h,i)perylene	4.060	4.254	-4.8	63	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0