

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE011018\
 Data File : BE095070.D
 Acq On : 10 Jan 2018 14:07
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jan 10 15:13:36 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 10 13:00:41 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	-0.01
2	1,4-Dioxane	0.394	0.389	1.3	98	0.00
3	n-Nitrosodimethylamine	0.466	0.472	-1.3	106	0.00
4 S	2-Fluorophenol	0.621	0.583	6.1	103	-0.01
5 S	Phenol-d6	0.923	0.855	7.4	101	-0.01
6	bis(2-Chloroethyl)ether	0.901	0.927	-2.9	110	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	112	-0.03
8 S	Nitrobenzene-d5	0.328	0.317	3.4	113	-0.02
9	Naphthalene	4.697	4.496	4.3	108	-0.03
10	Hexachlorobutadiene	0.214	0.204	4.7	107	-0.03
11 SURR	2-Methylnaphthalene-d10	0.655	0.643	1.8	111	-0.03
12	2-Methylnaphthalene	1.172	0.761	35.1#	74	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	-0.03
14 S	2,4,6-Tribromophenol	0.093	0.074	20.4	94	-0.03
15 S	2-Fluorobiphenyl	1.772	1.770	0.1	108	-0.02
16	Acenaphthylene	2.138	2.013	5.8	109	-0.03
17	Acenaphthene	1.418	1.361	4.0	107	-0.02
18	Fluorene	1.762	1.694	3.9	108	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	-0.03
20	4-Bromophenyl-phenylether	0.234	0.228	2.6	111	-0.03
21	Hexachlorobenzene	0.239	0.222	7.1	104	-0.03
22	Pentachlorophenol	0.046	0.033	28.3#	83	-0.02
23	Phenanthrene	1.246	1.180	5.3	108	-0.03
24	Anthracene	1.279	1.190	7.0	109	-0.03
25 SURR	Fluoranthene-d10	5.189	4.772	8.0	107	-0.04
26	Fluoranthene	1.538	1.404	8.7	104	-0.04
27 I	Chrysene-d12	1.000	1.000	0.0	97	-0.03
28	Pyrene	1.355	1.600	-18.1	111	-0.03
29 S	Terphenyl-d14	0.753	0.798	-6.0	103	-0.03
30	Benzo(a)anthracene	1.169	1.158	0.9	98	-0.05
31	Chrysene	1.316	1.401	-6.5	104	-0.05
32	Bis(2-ethylhexyl)phthalate	1.489	1.387	6.9	102	-0.05
33	Indeno(1,2,3-cd)pyrene	1.071	1.052	1.8	94	-0.09
34 I	Perylene-d12	1.000	1.000	0.0	104	-0.06
35	Benzo(b)fluoranthene	1.469	1.392	5.2	99	-0.05
36	Benzo(k)fluoranthene	1.488	1.441	3.2	106	-0.06
37 C	Benzo(a)pyrene	1.303	1.263	3.1	102	-0.06
38	Dibenzo(a,h)anthracene	1.209	0.963	20.3	84	-0.09
39	Benzo(g,h,i)perylene	1.203	1.100	8.6	97	-0.11

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE011018\
Data File : BE095070.D
Acq On : 10 Jan 2018 14:07
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jan 10 15:13:36 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 10 13:00:41 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		