

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011117\
 Data File : BE092637.D
 Acq On : 11 Jan 2017 11:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 11 16:37:38 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.733	0.796	-8.6	121	0.00
3	n-Nitrosodimethylamine	0.613	0.491	19.9	114	0.01
4 S	2-Fluorophenol	0.761	0.589	22.6	99	0.00
5 S	Phenol-d6	1.090	0.912	16.3	116	0.00
6	bis(2-Chloroethyl)ether	1.132	0.927	18.1	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.280	0.237	15.4	114	0.00
9	Naphthalene	1.026	0.986	3.9	111	0.00
10	Hexachlorobutadiene	0.156	0.148	5.1	108	0.00
11	2-Methylnaphthalene	0.653	0.583	10.7	107	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
13 S	2,4,6-Tribromophenol	0.113	0.082	27.4#	99	0.00
14 S	2-Fluorobiphenyl	1.196	1.080	9.7	108	0.00
15	Acenaphthylene	6.899	6.061	12.1	107	0.00
16	Acenaphthene	1.137	1.029	9.5	106	0.00
17	Fluorene	1.397	1.271	9.0	109	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
19	Hexachlorobenzene	0.214	0.187	12.6	106	0.00
20	Pentachlorophenol	0.036	0.026	27.8#	115	0.00
21	Phenanthrene	1.261	1.124	10.9	113	0.00
22	Anthracene	1.152	0.962	16.5	110	0.00
23	Fluoranthene	5.447	4.633	14.9	108	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
25	Pyrene	4.403	3.835	12.9	108	0.00
26 S	Terphenyl-d14	0.613	0.554	9.6	110	0.00
27	Benzo(a)anthracene	4.490	3.994	11.0	113	0.00
28	Chrysene	5.204	5.127	1.5	116	0.00
29	Bis(2-ethylhexyl)phthalate	0.486	0.369	24.1	117	0.00
30	Indeno(1,2,3-cd)pyrene	4.702	3.859	17.9	104	0.00
31 I	Perylene-d12	1.000	1.000	0.0	108	0.00
32	Benzo(b)fluoranthene	1.150	1.000	13.0	107	0.00
33	Benzo(k)fluoranthene	1.282	1.152	10.1	111	0.00
34 C	Benzo(a)pyrene	1.110	0.942	15.1	109	0.00
35	Dibenzo(a,h)anthracene	1.054	0.845	19.8	103	0.00
36	Benzo(g,h,i)perylene	4.515	3.849	14.8	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0