

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011117\
 Data File : BE092637.D
 Acq On : 11 Jan 2017 11:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 11 16:37:38 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	110	0.00
2	1,4-Dioxane	0.400	0.398	0.5	121	0.00
3	n-Nitrosodimethylamine	0.400	0.424	-6.0	114	0.01
4 S	2-Fluorophenol	0.400	0.386	3.5	99	0.00
5 S	Phenol-d6	0.400	0.414	-3.5	116	0.00
6	bis(2-Chloroethyl)ether	0.400	0.358	10.5	112	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.444	-11.0	114	0.00
9	Naphthalene	0.400	0.384	4.0	111	0.00
10	Hexachlorobutadiene	0.400	0.379	5.3	108	0.00
11	2-Methylnaphthalene	0.400	0.357	10.8	107	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	109	0.00
13 S	2,4,6-Tribromophenol	0.400	0.396	1.0	99	0.00
14 S	2-Fluorobiphenyl	0.400	0.361	9.8	108	0.00
15	Acenaphthylene	0.400	0.351	12.3	107	0.00
16	Acenaphthene	0.400	0.362	9.5	106	0.00
17	Fluorene	0.400	0.364	9.0	109	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	112	0.00
19	Hexachlorobenzene	0.400	0.351	12.3	106	0.00
20	Pentachlorophenol	0.400	0.371	7.3	115	0.00
21	Phenanthrene	0.400	0.356	11.0	113	0.00
22	Anthracene	0.400	0.334	16.5	110	0.00
23	Fluoranthene	0.400	0.340	15.0	108	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	109	0.00
25	Pyrene	0.400	0.348	13.0	108	0.00
26 S	Terphenyl-d14	0.400	0.361	9.8	110	0.00
27	Benzo(a)anthracene	0.400	0.356	11.0	113	0.00
28	Chrysene	0.400	0.394	1.5	116	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.393	1.8	117	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.328	18.0	104	0.00
31 I	Perylene-d12	5.000	5.000	0.0	108	0.00
32	Benzo(b)fluoranthene	0.400	0.348	13.0	107	0.00
33	Benzo(k)fluoranthene	0.400	0.359	10.3	111	0.00
34 C	Benzo(a)pyrene	0.400	0.339	15.3	109	0.00
35	Dibenzo(a,h)anthracene	0.400	0.321	19.8	103	0.00
36	Benzo(g,h,i)perylene	0.400	0.341	14.8	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0