

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011117\
 Data File : BE092646.D
 Acq On : 11 Jan 2017 18:47
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 12 00:21:54 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.733	0.556	24.1	89	0.00
3	n-Nitrosodimethylamine	0.613	0.549	10.4	134	0.00
4 S	2-Fluorophenol	0.761	0.774	-1.7	136	0.00
5 S	Phenol-d6	1.090	1.041	4.5	140	-0.02
6	bis(2-Chloroethyl)ether	1.132	1.205	-6.4	154#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.280	0.217	22.5	113	0.00
9	Naphthalene	1.026	0.963	6.1	118	0.00
10	Hexachlorobutadiene	0.156	0.148	5.1	117	0.00
11	2-Methylnaphthalene	0.653	0.586	10.3	116	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
13 S	2,4,6-Tribromophenol	0.113	0.095	15.9	122	0.00
14 S	2-Fluorobiphenyl	1.196	1.108	7.4	117	0.00
15	Acenaphthylene	6.899	6.263	9.2	118	0.00
16	Acenaphthene	1.137	1.060	6.8	116	0.00
17	Fluorene	1.397	1.302	6.8	119	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	115	-0.02
19	Hexachlorobenzene	0.214	0.206	3.7	120	0.00
20	Pentachlorophenol	0.036	0.035	2.8	162#	0.00
21	Phenanthrene	1.261	1.172	7.1	122	0.00
22	Anthracene	1.152	0.964	16.3	114	0.00
23	Fluoranthene	5.447	4.849	11.0	117	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
25	Pyrene	4.403	4.007	9.0	116	0.00
26 S	Terphenyl-d14	0.613	0.587	4.2	121	-0.02
27	Benzo(a)anthracene	4.490	4.167	7.2	122	0.00
28	Chrysene	5.204	4.884	6.1	114	0.00
29	Bis(2-ethylhexyl)phthalate	0.486	0.449	7.6	146	0.00
30	Indeno(1,2,3-cd)pyrene	4.702	4.341	7.7	121	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	117	0.00
32	Benzo(b)fluoranthene	1.150	1.054	8.3	122	0.00
33	Benzo(k)fluoranthene	1.282	1.133	11.6	118	0.00
34 C	Benzo(a)pyrene	1.110	0.988	11.0	124	0.00
35	Dibenzo(a,h)anthracene	1.054	0.915	13.2	120	0.00
36	Benzo(g,h,i)perylene	4.515	4.084	9.5	120	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0