

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011117\
 Data File : BE092646.D
 Acq On : 11 Jan 2017 18:47
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 12 00:21:54 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	115	0.00
2	1,4-Dioxane	0.400	0.223	44.3#	89	0.00
3	n-Nitrosodimethylamine	0.400	0.450	-12.5	134	0.00
4 S	2-Fluorophenol	0.400	0.462	-15.5	136	0.00
5 S	Phenol-d6	0.400	0.449	-12.2	140	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.433	-8.2	154	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.400	0.422	-5.5	113	0.00
9	Naphthalene	0.400	0.375	6.3	118	0.00
10	Hexachlorobutadiene	0.400	0.380	5.0	117	0.00
11	2-Methylnaphthalene	0.400	0.359	10.3	116	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	115	0.00
13 S	2,4,6-Tribromophenol	0.400	0.430	-7.5	122	0.00
14 S	2-Fluorobiphenyl	0.400	0.371	7.3	117	0.00
15	Acenaphthylene	0.400	0.363	9.3	118	0.00
16	Acenaphthene	0.400	0.373	6.8	116	0.00
17	Fluorene	0.400	0.373	6.8	119	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	115	-0.02
19	Hexachlorobenzene	0.400	0.385	3.8	120	0.00
20	Pentachlorophenol	0.400	0.495	-23.7	162	0.00
21	Phenanthrene	0.400	0.372	7.0	122	0.00
22	Anthracene	0.400	0.335	16.3	114	0.00
23	Fluoranthene	0.400	0.356	11.0	117	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	112	0.00
25	Pyrene	0.400	0.364	9.0	116	0.00
26 S	Terphenyl-d14	0.400	0.383	4.3	121	-0.02
27	Benzo(a)anthracene	0.400	0.371	7.3	122	0.00
28	Chrysene	0.400	0.375	6.3	114	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.461	-15.3	146	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	121	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	117	0.00
32	Benzo(b)fluoranthene	0.400	0.366	8.5	122	0.00
33	Benzo(k)fluoranthene	0.400	0.354	11.5	118	0.00
34 C	Benzo(a)pyrene	0.400	0.356	11.0	124	0.00
35	Dibenzo(a,h)anthracene	0.400	0.347	13.3	120	0.00
36	Benzo(g,h,i)perylene	0.400	0.362	9.5	120	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0