

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011117\
 Data File : BE092637.D
 Acq On : 11 Jan 2017 11:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

Sohil
 1/12/2017 8:30:43 AM

Quant Time: Jan 11 16:37:38 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	8445	5.00	ng	0.00
7) Naphthalene-d8	10.93	136	40805	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	29573	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	65161	5.00	ng	0.00
24) Chrysene-d12	21.72	240	82688	5.00	ng	0.00
31) Perylene-d12	24.06	264	74368	5.00	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.69	112	398	0.39	ng	0.00
5) Phenol-d6	7.38	99	616	0.41	ng	0.00
8) Nitrobenzene-d5	9.36	82	773m	0.44	ng	0.00
13) 2,4,6-Tribromophenol	16.31	330	194	0.40	ng	0.00
14) 2-Fluorobiphenyl	13.43	172	2555	0.36	ng	0.00
26) Terphenyl-d14	20.17	244	3665	0.36	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.39	88	538	0.40	ng	88
3) n-Nitrosodimethylamine	3.80	42	332	0.42	ng	# 93
6) bis(2-Chloroethyl)ether	7.59	93	626	0.36	ng	# 79
9) Naphthalene	10.97	128	3219	0.38	ng	97
10) Hexachlorobutadiene	11.26	225	482	0.38	ng	99
11) 2-Methylnaphthalene	12.64	142	1902	0.36	ng	98
15) Acenaphthylene	14.51	152	14339	0.35	ng	100
16) Acenaphthene	14.85	154	2435	0.36	ng	96
17) Fluorene	15.86	166	3006	0.36	ng	99
19) Hexachlorobenzene	16.87	284	977m	0.35	ng	
20) Pentachlorophenol	17.23	266	134	0.37	ng	# 88
21) Phenanthrene	17.58	178	5860	0.36	ng	# 95
22) Anthracene	17.70	178	5016	0.33	ng	99
23) Fluoranthene	19.59	202	24152	0.34	ng	99
25) Pyrene	19.95	202	25369	0.35	ng	100
27) Benzo(a)anthracene	21.70	228	26420m	0.36	ng	
28) Chrysene	21.75	228	33915m	0.39	ng	
29) Bis(2-ethylhexyl)phthalate	21.61	149	2444	0.39	ng	# 75
30) Indeno(1,2,3-cd)pyrene	26.56	276	25529	0.33	ng	99
32) Benzo(b)fluoranthene	23.35	252	5949	0.35	ng	98
33) Benzo(k)fluoranthene	23.39	252	6853	0.36	ng	97
34) Benzo(a)pyrene	23.96	252	5603	0.34	ng	99
35) Dibenzo(a,h)anthracene	26.58	278	5026	0.32	ng	# 95
36) Benzo(g,h,i)perylene	27.33	276	22901	0.34	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE011117\
Data File : BE092637.D
Acq On : 11 Jan 2017 11:59
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM

Quant Time: Jan 11 16:37:38 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 19 16:16:32 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4

Tot Ion:152 Resp: 8445

Ion Ratio Lower Upper

152 100


150 172.0 106.0 159.0#

115 77.3 31.2 46.8#




Manual Integrations
APPROVED

Sohil

1/12/2017 8:30:43 AM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.40 ng

RT: 3.39 min Scan# 69

Delta R.T. 0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59


Tgt Ion: 88 Resp: 538

Ion Ratio Lower Upper



88 100


58 67.8 63.6 95.4

43 29.4 28.0 42.0



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0





#3





n-Nitrosodimethylamine
Concen: 0.42 ng
RT: 3.80 min Scan# 105
Delta R.T. 0.01 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
42	100		
74	100.9	86.0	129.0
44	0.0	5.1	7.7#

Manual Integrations
APPROVED




Sohil
1/12/2017 8:30:43 AM




#4

2-Fluorophenol
Concen: 0.39 ng
RT: 5.69 min Scan# 367
Delta R.T. 0.01 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
112	100		
64	75.9	53.5	80.3
63	30.2	27.9	41.9





#5

Phenol-d6

Concen: 0.41 ng

RT: 7.38 min Scan# 542

Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	99	Resp:	616
Ion	Ratio	Lower	Upper
99	100		
42	0.0	16.0	24.0#
71	0.0	30.6	45.8#

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM



#6

bis(2-Chloroethyl)ether

Concen: 0.36 ng




RT: 7.59 min Scan# 551


Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Tgt Ion:	93	Resp:	626
Ion	Ratio	Lower	Upper
93	100		
63	64.7	71.6	107.4#
95	31.9	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.93 min Scan# 707

Delta R.T. 0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	11.8	8.2	12.4	
54	7.1	9.6	14.4	
68	5.0	6.7	10.1	

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM



#8

Nitrobenzene-d5

Concen: 0.44 ng m




RT: 9.36 min Scan# 628


Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	47.7	39.0	58.4	
54	76.2	44.8	67.2	





#9

Naphthalene

Concen: 0.38 ng

RT: 10.97 min Scan# 709

Delta R.T. 0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	128	Resp:	3219
Ion	Ratio	Lower	Upper
128	100		
129	14.2	11.2	16.8
127	16.7	11.6	17.4

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM



#10

Hexachlorobutadiene

Concen: 0.38 ng




RT: 11.26 min Scan# 724


Delta R.T. 0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Tgt Ion:	225	Resp:	482
Ion	Ratio	Lower	Upper
225	100		
223	64.3	50.6	76.0
227	63.9	51.4	77.0





#11

2-Methylnaphthalene

Concen: 0.36 ng

RT: 12.64 min Scan# 815

Delta R.T. 0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E

ClientSampled :





SSTDCCC0.4

Tot Ion:142 Resp: 1902

Ion	Ratio	Lower	Upper
142	100		
141	90.2	73.7	110.5
115	41.9	34.0	51.0

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.80 min Scan# 979




Delta R.T. -0.00 min


Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Tgt Ion:164 Resp: 29573

Ion	Ratio	Lower	Upper
164	100		
162	92.3	71.4	107.0
160	37.9	27.4	41.2









#13
 2,4,6-Tribromophenol
 Concen: 0.40 ng
 RT: 16.31 min Scan# 1088
 Delta R.T. 0.00 min
 Lab File: BE092637.D
 Acq: 11 Jan 2017 11:59

Instrument :
 BNA_E
 ClientSampled :
 SSTDCCC0.4

Tot Ion:330 Resp: 194
 Ion Ratio Lower Upper
 330 100
 332 97.9 75.4 113.0
 141 42.8 31.0 46.6




Manual Integrations
 APPROVED


Sohil
 1/12/2017 8:30:43 AM



#14
 2-Fluorobiphenyl
 Concen: 0.36 ng
 RT: 13.43 min Scan# 884
 Delta R.T. 0.00 min
 Lab File: BE092637.D
 Acq: 11 Jan 2017 11:59

Tgt Ion:172 Resp: 2555
 Ion Ratio Lower Upper
 172 100
 171 36.7 30.1 45.1
 170 26.3 21.8 32.6









#15
Acenaphthylene
Concen: 0.35 ng
RT: 14.51 min Scan# 962
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.2	10.4	15.6




Manual Integrations
APPROVED


Sohil
1/12/2017 8:30:43 AM



#16
Acenaphthene
Concen: 0.36 ng
RT: 14.85 min Scan# 982
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
154	100		
153	441.3	350.8	526.2
152	232.0	171.1	256.7









#17
Fluorene
Concen: 0.36 ng
RT: 15.86 min Scan# 1049
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
166	100		
165	99.3	78.2	117.4
167	13.3	11.0	16.4




Manual Integrations
APPROVED


Sohil
1/12/2017 8:30:43 AM



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.56 min Scan# 1153
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
188	100		
94	4.0	3.2	4.8
80	3.6	2.6	3.8






#19
Hexachlorobenzene
Concen: 0.35 ng/ml
RT: 16.87 min Scan# 1123
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4




Tot Ion	Ratio	Resp	Lower	Upper
284	100	977		
142	0.0	29.1	43.7#	
249	45.0	27.9	41.9#	

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM






Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E




#20
Pentachlorophenol
Concen: 0.37 ng/ml
RT: 17.23 min Scan# 1139
Delta R.T. 0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	134		
268	84.2	62.5	93.7	
264	86.8	57.2	85.8#	

Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E









#21
Phenanthrene
Concen: 0.36 ng
RT: 17.58 min Scan# 1154
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.8	23.8
179	15.9	16.0	24.0#




Manual Integrations
APPROVED


Sohil
1/12/2017 8:30:43 AM



#22
Anthracene
Concen: 0.33 ng
RT: 17.70 min Scan# 1159
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	15.9	12.2	18.2





#23

Fluoranthene

Concen: 0.34 ng

RT: 19.59 min Scan# 1265

Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
202	100		
101	2.9	2.2	3.4
203	17.5	13.7	20.5

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.72 min Scan# 1374


Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
240	100		
120	1.9	2.6	4.0#
236	25.0	24.6	37.0









#25
Pvrene
Concen: 0.35 ng
RT: 19.95 min Scan# 1286
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
202	100		
200	5.3	4.2	6.4
203	17.6	13.9	20.9




Manual Integrations
APPROVED


Sohil
1/12/2017 8:30:43 AM



#26
Terphenyl-d14
Concen: 0.36 ng
RT: 20.17 min Scan# 1299
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.2	6.7	10.1#
122	7.7	3.6	5.4#









#27
Benzo(a)anthracene
Concen: 0.36 ng m
RT: 21.70 min Scan# 1373
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
228	100		
226	23.7	23.6	35.4
229	20.9	13.3	19.9




Manual Integrations
APPROVED


Sohil
1/12/2017 8:30:43 AM



#28
Chrysene
Concen: 0.39 ng m
RT: 21.75 min Scan# 1375
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.9	36.3	54.5
229	16.9	10.6	16.0





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.39 ng

RT: 21.61 min Scan# 1369

Delta R.T. 0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4




Tot Ion:	149	Resp:	2444
Ion Ratio	Lower	Upper	
149	100		
167	22.8	16.0	24.0
279	0.0	16.0	24.0#

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.33 ng

RT: 26.56 min Scan# 1714




Delta R.T. -0.00 min


Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Tgt Ion:	276	Resp:	25529
Ion Ratio	Lower	Upper	
276	100		
138	26.5	20.3	30.5
277	24.8	19.7	29.5

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.06 min Scan# 1542
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
264	100		
260	26.1	20.1	30.1
265	20.8	17.9	26.9




Manual Integrations
APPROVED


Sohil
1/12/2017 8:30:43 AM



#32
Benzo(b)fluoranthene
Concen: 0.35 ng
RT: 23.35 min Scan# 1480
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.0	18.7	28.1
125	12.3	10.5	15.7









#33
Benzo(k)fluoranthene
Concen: 0.36 ng
RT: 23.39 min Scan# 1484
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.7	20.3	30.5
125	12.3	9.6	14.4




Manual Integrations
APPROVED


Sohil
1/12/2017 8:30:43 AM



#34
Benzo(a)pyrene
Concen: 0.34 ng
RT: 23.96 min Scan# 1533
Delta R.T. -0.00 min
Lab File: BE092637.D
Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.2	19.0	28.6
125	14.6	11.8	17.8





#35

Dibenzo(a,h)anthracene

Concen: 0.32 ng

RT: 26.58 min Scan# 1715

Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
278	100		
139	22.2	13.8	20.8#
279	27.5	21.4	32.2

Manual Integrations
APPROVED

Sohil
1/12/2017 8:30:43 AM



#36

Benzo(g,h,i)perylene

Concen: 0.34 ng

RT: 27.33 min Scan# 1759

Delta R.T. -0.00 min

Lab File: BE092637.D

Acq: 11 Jan 2017 11:59

Tgt Ion	Ratio	Lower	Upper
276	100		
277	27.9	19.8	29.8
138	24.6	19.8	29.6

