

Data Path : Z:\HPCHEM1\BNA_E\Data\BE011416\
 Data File : BE091354.D
 Acq On : 14 Jan 2016 16:56
 Operator : UM/IZ
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC001

Quant Time: Jan 15 02:58:03 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 14 23:29:02 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	22857	5.00	ng	0.00
4) Naphthalene-d8	9.55	136	103102	5.00	ng	0.00
8) Acenaphthene-d10	13.35	164	47464	5.00	ng	0.00
14) Phenanthrene-d10	16.08	188	135814	5.00	ng	0.00
18) Chrysene-d12	20.21	240	135533	5.00	ng	0.00
24) Perylene-d12	22.10	264	111688	5.00	ng	0.00
System Monitoring Compounds						
2) 2-Fluorophenol	4.95	112	2971	1.19	ng	0.00
3) Phenol-d6	6.57	99	4877	1.18	ng	0.00
5) Nitrobenzene-d5	8.17	82	6109	1.02	ng	0.00
9) 2,4,6-Tribromophenol	14.89	330	810	0.96	ng	0.00
10) 2-Fluorobiphenyl	12.02	172	10998	1.07	ng	0.00
20) Terphenyl-d14	18.62	244	12512	0.92	ng	0.00
Target Compounds						Qvalue
6) Naphthalene	9.59	128	22075	1.00	ng	99
7) 2-Methylnaphthalene	11.13	142	8057	0.92	ng	99
11) Acenaphthylene	13.11	152	63749	0.95	ng	100
12) Acenaphthene	13.42	154	13688	1.02	ng	92
13) Fluorene	14.41	166	15945	0.96	ng	99
15) Phenanthrene	16.12	178	34536	1.03	ng	100
16) Anthracene	16.20	178	24000	0.94	ng	100
17) Fluoranthene	18.10	202	20999	0.84	ng	99
19) Pyrene	18.48	202	19333	0.93	ng	100
21) Benzo(a)anthracene	20.19	228	16228	0.90	ng	99
22) Chrysene	20.25	228	34580	1.05	ng	100
23) Indeno(1,2,3-cd)pyrene	23.71	276	15282	0.78	ng	96
25) Benzo(b)fluoranthene	21.55	252	19303	0.93	ng	99
26) Benzo(k)fluoranthene	21.58	252	24485	0.99	ng	99
27) Benzo(a)pyrene	22.01	252	16000	0.81	ng	100
28) Dibenzo(a,h)anthracene	23.66	278	10234	0.81	ng	98
29) Benzo(g,h,i)perylene	24.25	276	16895	0.92	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\Data\BE011416\
Data File : BE091354.D
Acq On : 14 Jan 2016 16:56
Operator : UM/IZ
Sample : SSTDICCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Quant Time: Jan 15 02:58:03 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 14 23:29:02 2016
Response via : Initial Calibration

