

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE011416\
 Data File : BE091363.D
 Acq On : 14 Jan 2016 23:00
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE011416

Quant Time: Jan 15 15:15:23 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 14 23:29:02 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	80	0.00
2 S	2-Fluorophenol	1.000	1.309	-30.9#	108	0.00
3 S	Phenol-d6	1.000	1.275	-27.5#	130	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	84	0.00
5 S	Nitrobenzene-d5	1.000	1.046	-4.6	88	0.00
6	Naphthalene	1.000	1.057	-5.7	87	0.00
7	2-Methylnaphthalene	1.000	0.955	4.5	100	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	89	0.00
9 S	2,4,6-Tribromophenol	1.000	0.982	1.8	103	0.00
10 S	2-Fluorobiphenyl	1.000	1.078	-7.8	102	0.00
11	Acenaphthylene	1.000	0.961	3.9	103	0.00
12	Acenaphthene	1.000	1.031	-3.1	93	0.00
13	Fluorene	1.000	1.025	-2.5	104	0.00
14 I	Phenanthrene-d10	5.000	5.000	0.0	91	0.00
15	Phenanthrene	1.000	1.014	-1.4	90	0.00
16	Anthracene	1.000	0.959	4.1	99	0.00
17	Fluoranthene	1.000	0.812	18.8	69	0.00
18 I	Chrysene-d12	5.000	5.000	0.0	82	0.00
19	Pyrene	1.000	0.906	9.4	91	0.00
20 S	Terphenyl-d14	1.000	0.962	3.8	96	0.00
21	Benzo(a)anthracene	1.000	0.901	9.9	93	0.00
22	Chrysene	1.000	1.062	-6.2	87	0.00
23	Indeno(1,2,3-cd)pyrene	1.000	0.930	7.0	89	0.00
24 I	Perylene-d12	5.000	5.000	0.0	75	0.00
25	Benzo(b)fluoranthene	1.000	0.994	0.6	96	0.00
26	Benzo(k)fluoranthene	1.000	0.977	2.3	84	0.00
27 C	Benzo(a)pyrene	1.000	0.860	14.0	68	0.00
28	Dibenzo(a,h)anthracene	1.000	0.967	3.3	96	0.00
29	Benzo(g,h,i)perylene	1.000	0.945	5.5	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0