

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE011416\
 Data File : BE091363.D
 Acq On : 14 Jan 2016 23:00
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE011416

Quant Time: Jan 15 15:15:23 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 14 23:29:02 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2 S	2-Fluorophenol	0.533	0.723	-35.6#	108	0.00
3 S	Phenol-d6	0.852	1.170	-37.3#	130	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	84	0.00
5 S	Nitrobenzene-d5	0.290	0.303	-4.5	88	0.00
6	Naphthalene	1.067	1.127	-5.6	87	0.00
7	2-Methylnaphthalene	0.417	0.413	1.0	100	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
9 S	2,4,6-Tribromophenol	0.103	0.090	12.6	103	0.00
10 S	2-Fluorobiphenyl	1.081	1.165	-7.8	102	0.00
11	Acenaphthylene	6.555	6.803	-3.8	103	0.00
12	Acenaphthene	1.419	1.463	-3.1	93	0.00
13	Fluorene	1.575	1.818	-15.4	104	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
15	Phenanthrene	1.237	1.255	-1.5	90	0.00
16	Anthracene	0.867	0.904	-4.3	99	0.00
17	Fluoranthene	0.893	0.741	17.0	69	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
19	Pyrene	0.764	0.685	10.3	91	0.00
20 S	Terphenyl-d14	0.474	0.491	-3.6	96	0.00
21	Benzo(a)anthracene	0.696	0.606	12.9	93	0.00
22	Chrysene	1.214	1.289	-6.2	87	0.00
23	Indeno(1,2,3-cd)pyrene	0.793	0.758	4.4	89	0.00
24 I	Perylene-d12	1.000	1.000	0.0	75	0.00
25	Benzo(b)fluoranthene	0.839	0.952	-13.5	96	0.00
26	Benzo(k)fluoranthene	0.981	1.078	-9.9	84	0.00
27 C	Benzo(a)pyrene	0.857	0.775	9.6	68	0.00
28	Dibenzo(a,h)anthracene	0.637	0.653	-2.5	96	0.00
29	Benzo(g,h,i)perylene	0.811	0.795	2.0	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0