

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE011416\
 Data File : BE091370.D
 Acq On : 15 Jan 2016 3:17
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 15 15:22:36 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 14 23:29:02 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2 S	2-Fluorophenol	0.533	0.895	-67.9#	205#	-0.02
3 S	Phenol-d6	0.852	1.188	-39.4#	202#	-0.01
4 I	Naphthalene-d8	1.000	1.000	0.0	132	0.00
5 S	Nitrobenzene-d5	0.290	0.307	-5.9	140	0.00
6	Naphthalene	1.067	1.088	-2.0	132	0.00
7	2-Methylnaphthalene	0.417	0.471	-12.9	178#	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
9 S	2,4,6-Tribromophenol	0.103	0.100	2.9	182#	0.00
10 S	2-Fluorobiphenyl	1.081	1.169	-8.1	164#	0.00
11	Acenaphthylene	6.555	7.131	-8.8	173#	0.00
12	Acenaphthene	1.419	1.363	3.9	138	0.00
13	Fluorene	1.575	1.717	-9.0	157#	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
15	Phenanthrene	1.237	1.253	-1.3	123	-0.01
16	Anthracene	0.867	1.003	-15.7	152#	0.00
17	Fluoranthene	0.893	1.296	-45.1#	167#	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
19	Pyrene	0.764	0.948	-24.1	201#	0.00
20 S	Terphenyl-d14	0.474	0.594	-25.3#	185#	0.00
21	Benzo(a)anthracene	0.696	0.965	-38.6#	236#	0.00
22	Chrysene	1.214	1.274	-4.9	137	0.00
23	Indeno(1,2,3-cd)pyrene	0.793	1.110	-40.0#	209#	0.00
24 I	Perylene-d12	1.000	1.000	0.0	141	0.00
25	Benzo(b)fluoranthene	0.839	1.183	-41.0#	224#	0.00
26	Benzo(k)fluoranthene	0.981	1.177	-20.0	171#	0.00
27 C	Benzo(a)pyrene	0.857	1.062	-23.9#	174#	0.00
28	Dibenzo(a,h)anthracene	0.637	0.958	-50.4#	262#	0.00
29	Benzo(g,h,i)perylene	0.811	1.059	-30.6#	210#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1