

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011516\
 Data File : BE091372.D
 Acq On : 15 Jan 2016 9:38
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 16 05:03:17 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 15 15:14:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2 S	2-Fluorophenol	0.533	0.843	-58.2#	145	-0.01
3 S	Phenol-d6	0.852	1.054	-23.7	135	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
5 S	Nitrobenzene-d5	0.290	0.289	0.3	98	0.00
6	Naphthalene	1.067	1.079	-1.1	97	0.00
7	2-Methylnaphthalene	0.417	0.452	-8.4	127	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
9 S	2,4,6-Tribromophenol	0.103	0.100	2.9	134	0.00
10 S	2-Fluorobiphenyl	1.081	1.167	-8.0	121	0.00
11	Acenaphthylene	6.555	7.125	-8.7	127	0.00
12	Acenaphthene	1.419	1.389	2.1	103	0.00
13	Fluorene	1.575	1.698	-7.8	114	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
15	Phenanthrene	1.237	1.284	-3.8	94	-0.01
16	Anthracene	0.867	0.994	-14.6	112	0.00
17	Fluoranthene	0.893	1.041	-16.6	100	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
19	Pyrene	0.764	0.808	-5.8	126	0.00
20 S	Terphenyl-d14	0.474	0.544	-14.8	124	0.00
21	Benzo(a)anthracene	0.696	0.768	-10.3	138	0.00
22	Chrysene	1.214	1.231	-1.4	97	0.00
23	Indeno(1,2,3-cd)pyrene	0.793	0.823	-3.8	113	0.00
24 I	Perylene-d12	1.000	1.000	0.0	98	0.00
25	Benzo(b)fluoranthene	0.839	0.999	-19.1	131	0.00
26	Benzo(k)fluoranthene	0.981	1.186	-20.9	120	0.00
27 C	Benzo(a)pyrene	0.857	0.914	-6.7	104	0.00
28	Dibenzo(a,h)anthracene	0.637	0.795	-24.8	151#	0.00
29	Benzo(g,h,i)perylene	0.811	0.947	-16.8	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0