

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011516\
 Data File : BE091372.D
 Acq On : 15 Jan 2016 9:38
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 16 05:03:17 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 15 15:14:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	91	0.00
2 S	2-Fluorophenol	1.000	1.500	-50.0#	145	-0.01
3 S	Phenol-d6	1.000	1.171	-17.1	135	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	97	0.00
5 S	Nitrobenzene-d5	1.000	0.982	1.8	98	0.00
6	Naphthalene	1.000	1.012	-1.2	97	0.00
7	2-Methylnaphthalene	1.000	1.009	-0.9	127	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	105	0.00
9 S	2,4,6-Tribromophenol	1.000	1.032	-3.2	134	0.00
10 S	2-Fluorobiphenyl	1.000	1.029	-2.9	121	0.00
11	Acenaphthylene	1.000	0.994	0.6	127	0.00
12	Acenaphthene	1.000	0.979	2.1	103	0.00
13	Fluorene	1.000	0.967	3.3	114	0.00
14 I	Phenanthrene-d10	5.000	5.000	0.0	93	0.00
15	Phenanthrene	1.000	1.037	-3.7	94	-0.01
16	Anthracene	1.000	1.029	-2.9	112	0.00
17	Fluoranthene	1.000	1.030	-3.0	100	0.00
18 I	Chrysene-d12	5.000	5.000	0.0	95	0.00
19	Pyrene	1.000	0.998	0.2	126	0.00
20 S	Terphenyl-d14	1.000	1.031	-3.1	124	0.00
21	Benzo(a)anthracene	1.000	1.040	-4.0	138	0.00
22	Chrysene	1.000	1.014	-1.4	97	0.00
23	Indeno(1,2,3-cd)pyrene	1.000	0.980	2.0	113	0.00
24 I	Perylene-d12	5.000	5.000	0.0	98	0.00
25	Benzo(b)fluoranthene	1.000	1.029	-2.9	131	0.00
26	Benzo(k)fluoranthene	1.000	1.049	-4.9	120	0.00
27 C	Benzo(a)pyrene	1.000	0.967	3.3	104	0.00
28	Dibenzo(a,h)anthracene	1.000	1.083	-8.3	151	0.00
29	Benzo(g,h,i)perylene	1.000	1.060	-6.0	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0