

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011516\
 Data File : BE091374.D
 Acq On : 15 Jan 2016 11:11
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 16 05:04:15 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 15 15:14:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2 S	2-Fluorophenol	0.533	0.718	-34.7#	115	0.00
3 S	Phenol-d6	0.852	1.194	-40.1#	142	-0.01
4 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
5 S	Nitrobenzene-d5	0.290	0.278	4.1	86	0.00
6	Naphthalene	1.067	1.073	-0.6	88	0.00
7	2-Methylnaphthalene	0.417	0.423	-1.4	108	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
9 S	2,4,6-Tribromophenol	0.103	0.090	12.6	107	0.00
10 S	2-Fluorobiphenyl	1.081	1.140	-5.5	105	0.00
11	Acenaphthylene	6.555	6.878	-4.9	110	0.00
12	Acenaphthene	1.419	1.420	-0.1	94	0.00
13	Fluorene	1.575	1.720	-9.2	103	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
15	Phenanthrene	1.237	1.240	-0.2	86	-0.01
16	Anthracene	0.867	0.943	-8.8	101	0.00
17	Fluoranthene	0.893	0.818	8.4	74	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
19	Pyrene	0.764	0.728	4.7	102	0.00
20 S	Terphenyl-d14	0.474	0.504	-6.3	104	0.00
21	Benzo(a)anthracene	0.696	0.646	7.2	105	0.00
22	Chrysene	1.214	1.223	-0.7	87	0.00
23	Indeno(1,2,3-cd)pyrene	0.793	0.731	7.8	91	0.00
24 I	Perylene-d12	1.000	1.000	0.0	85	0.00
25	Benzo(b)fluoranthene	0.839	0.915	-9.1	105	0.00
26	Benzo(k)fluoranthene	0.981	1.112	-13.4	98	0.00
27 C	Benzo(a)pyrene	0.857	0.811	5.4	80	0.00
28	Dibenzo(a,h)anthracene	0.637	0.681	-6.9	113	0.00
29	Benzo(g,h,i)perylene	0.811	0.827	-2.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0