

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011516\
 Data File : BE091374.D
 Acq On : 15 Jan 2016 11:11
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 16 05:04:15 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 15 15:14:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	85	0.00
2 S	2-Fluorophenol	1.000	1.301	-30.1#	115	0.00
3 S	Phenol-d6	1.000	1.297	-29.7#	142	-0.01
4 I	Naphthalene-d8	5.000	5.000	0.0	89	0.00
5 S	Nitrobenzene-d5	1.000	0.953	4.7	86	0.00
6	Naphthalene	1.000	1.006	-0.6	88	0.00
7	2-Methylnaphthalene	1.000	0.968	3.2	108	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	94	0.00
9 S	2,4,6-Tribromophenol	1.000	0.979	2.1	107	0.00
10 S	2-Fluorobiphenyl	1.000	1.010	-1.0	105	0.00
11	Acenaphthylene	1.000	0.969	3.1	110	0.00
12	Acenaphthene	1.000	1.001	-0.1	94	0.00
13	Fluorene	1.000	0.977	2.3	103	0.00
14 I	Phenanthrene-d10	5.000	5.000	0.0	88	0.00
15	Phenanthrene	1.000	1.002	-0.2	86	-0.01
16	Anthracene	1.000	0.989	1.1	101	0.00
17	Fluoranthene	1.000	0.868	13.2	74	0.00
18 I	Chrysene-d12	5.000	5.000	0.0	86	0.00
19	Pyrene	1.000	0.938	6.2	102	0.00
20 S	Terphenyl-d14	1.000	0.979	2.1	104	0.00
21	Benzo(a)anthracene	1.000	0.935	6.5	105	0.00
22	Chrysene	1.000	1.008	-0.8	87	0.00
23	Indeno(1,2,3-cd)pyrene	1.000	0.909	9.1	91	0.00
24 I	Perylene-d12	5.000	5.000	0.0	85	0.00
25	Benzo(b)fluoranthene	1.000	0.966	3.4	105	0.00
26	Benzo(k)fluoranthene	1.000	1.000	0.0	98	0.00
27 C	Benzo(a)pyrene	1.000	0.887	11.3	80	0.00
28	Dibenzo(a,h)anthracene	1.000	0.990	1.0	113	0.00
29	Benzo(g,h,i)perylene	1.000	0.969	3.1	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0