

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011520\
 Data File : BE101060.D
 Acq On : 15 Jan 2020 16:23
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 16 04:19:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	-0.01
2	1,4-Dioxane	1.020	1.027	-0.7	103	0.00
3	n-Nitrosodimethylamine	0.556	0.636	-14.4	127	-0.01
4 S	2-Fluorophenol	0.900	1.124	-24.9	120	-0.01
5 S	Phenol-d6	1.139	1.325	-16.3	125	-0.01
6	bis(2-Chloroethyl)ether	1.357	1.277	5.9	98	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
8 S	Nitrobenzene-d5	0.286	0.321	-12.2	117	0.00
9	Naphthalene	4.441	4.556	-2.6	94	0.00
10	Hexachlorobutadiene	0.188	0.198	-5.3	94	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.742	2.8	90	-0.02
12	2-Methylnaphthalene	0.672	0.640	4.8	90	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.113	0.119	-5.3	103	-0.01
15 S	2-Fluorobiphenyl	1.519	1.559	-2.6	86	-0.02
16	Acenaphthylene	1.629	1.730	-6.2	100	-0.02
17	Acenaphthene	1.166	1.192	-2.2	89	-0.02
18	Fluorene	1.474	1.426	3.3	84	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
20	4-Bromophenyl-phenylether	0.195	0.199	-2.1	85	-0.01
21	Hexachlorobenzene	0.214	0.227	-6.1	84	-0.01
22	Pentachlorophenol	0.053	0.055	-3.8	108	-0.01
23	Phenanthrene	1.092	1.044	4.4	77	0.00
24	Anthracene	1.031	1.075	-4.3	87	0.00
25 SURR	Fluoranthene-d10	5.555	5.410	2.6	80	-0.01
26	Fluoranthene	1.368	1.300	5.0	81	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	86	-0.01
28	Pvrene	1.489	1.462	1.8	83	0.00
29 S	Terphenyl-d14	0.982	0.928	5.5	80	0.00
30	Benzo(a)anthracene	1.141	1.163	-1.9	92	0.00
31	Chrysene	1.696	1.825	-7.6	90	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	2.675	-40.6#	140	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.830	-36.6#	121	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	116	-0.01
35	Benzo(b)fluoranthene	1.122	0.956	14.8	106	0.00
36	Benzo(k)fluoranthene	1.626	1.365	16.1	97	0.00
37 C	Benzo(a)pyrene	1.085	1.102	-1.6	121	0.00
38	Dibenzo(a,h)anthracene	0.998	0.990	0.8	124	-0.02
39	Benzo(a,h,i)perylene	1.227	1.234	-0.6	113	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011520\
Data File : BE101060.D
Acq On : 15 Jan 2020 16:23
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 16 04:19:09 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			