

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011520\
 Data File : BE101068.D
 Acq On : 15 Jan 2020 21:33
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 16 04:43:56 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.01
2	1,4-Dioxane	1.020	0.926	9.2	108	0.00
3	n-Nitrosodimethylamine	0.556	0.634	-14.0	146	-0.01
4 S	2-Fluorophenol	0.900	1.143	-27.0#	142	-0.01
5 S	Phenol-d6	1.139	1.358	-19.2	148	-0.01
6	bis(2-Chloroethyl)ether	1.357	1.324	2.4	118	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.286	0.308	-7.7	132	0.00
9	Naphthalene	4.441	4.774	-7.5	116	0.00
10	Hexachlorobutadiene	0.188	0.204	-8.5	115	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.699	8.4	100	-0.01
12	2-Methylnaphthalene	0.672	0.680	-1.2	113	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.113	0.112	0.9	116	-0.01
15 S	2-Fluorobiphenyl	1.519	1.547	-1.8	101	-0.01
16	Acenaphthylene	1.629	1.826	-12.1	125	-0.01
17	Acenaphthene	1.166	1.252	-7.4	111	-0.01
18	Fluorene	1.474	1.534	-4.1	107	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.195	0.197	-1.0	101	-0.01
21	Hexachlorobenzene	0.214	0.228	-6.5	101	-0.01
22	Pentachlorophenol	0.053	0.055	-3.8	130	-0.01
23	Phenanthrene	1.092	1.105	-1.2	98	0.00
24	Anthracene	1.031	1.124	-9.0	110	0.00
25 SURR	Fluoranthene-d10	5.555	5.373	3.3	95	-0.01
26	Fluoranthene	1.368	1.406	-2.8	104	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
28	Pvrene	1.489	1.543	-3.6	105	0.00
29 S	Terphenyl-d14	0.982	0.920	6.3	96	-0.01
30	Benzo(a)anthracene	1.141	1.254	-9.9	119	0.00
31	Chrysene	1.696	1.925	-13.5	115	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	2.776	-46.0#	175#	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.985	-48.1#	158#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	143	-0.01
35	Benzo(b)fluoranthene	1.122	1.006	10.3	137	-0.01
36	Benzo(k)fluoranthene	1.626	1.355	16.7	119	0.00
37 C	Benzo(a)pyrene	1.085	1.139	-5.0	155#	-0.01
38	Dibenzo(a,h)anthracene	0.998	1.110	-11.2	172#	-0.01
39	Benzo(a,h,i)perylene	1.227	1.271	-3.6	144	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011520\
Data File : BE101068.D
Acq On : 15 Jan 2020 21:33
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 16 04:43:56 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			