

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011617\
 Data File : BE092656.D
 Acq On : 16 Jan 2017 17:42
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE011617

Quant Time: Jan 17 10:21:48 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.775	0.782	-0.9	104	0.00
3	n-Nitrosodimethylamine	0.902	0.781	13.4	107	0.01
4 S	2-Fluorophenol	1.116	0.900	19.4	103	0.00
5 S	Phenol-d6	1.368	1.121	18.1	108	0.00
6	bis(2-Chloroethyl)ether	1.453	1.216	16.3	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.325	0.269	17.2	105	0.00
9	Naphthalene	1.044	1.000	4.2	109	0.00
10	Hexachlorobutadiene	0.154	0.148	3.9	106	0.00
11	2-Methylnaphthalene	0.648	0.581	10.3	106	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
13 S	2,4,6-Tribromophenol	0.126	0.106	15.9	114	0.00
14 S	2-Fluorobiphenyl	1.230	1.138	7.5	108	0.00
15	Acenaphthylene	6.890	6.148	10.8	109	0.00
16	Acenaphthene	1.105	1.044	5.5	108	0.00
17	Fluorene	1.387	1.255	9.5	109	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
19	Hexachlorobenzene	0.189	0.178	5.8	101	0.00
20	Pentachlorophenol	0.051	0.044	13.7	104	0.00
21	Phenanthrene	1.104	1.069	3.2	112	0.00
22	Anthracene	1.018	0.925	9.1	114	0.00
23	Fluoranthene	4.895	4.427	9.6	111	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
25	Pyrene	4.929	4.173	15.3	111	0.00
26 S	Terphenyl-d14	0.707	0.596	15.7	107	0.00
27	Benzo(a)anthracene	4.926	4.175	15.2	108	0.00
28	Chrysene	5.542	5.283	4.7	112	0.00
29	Bis(2-ethylhexyl)phthalate	0.913	0.536	41.3#	90	0.00
30	Indeno(1,2,3-cd)pyrene	5.987	4.931	17.6	103	0.00
31 I	Perylene-d12	1.000	1.000	0.0	106	0.00
32	Benzo(b)fluoranthene	1.210	0.994	17.9	95	-0.01
33	Benzo(k)fluoranthene	1.215	1.178	3.0	120	-0.01
34 C	Benzo(a)pyrene	1.136	0.992	12.7	107	-0.01
35	Dibenzo(a,h)anthracene	1.112	0.944	15.1	102	-0.02
36	Benzo(g,h,i)perylene	4.699	4.065	13.5	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0