

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011617\
 Data File : BE092654.D
 Acq On : 16 Jan 2017 16:28
 Operator : UM/SJ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

mohammad
 1/17/2017 4:40:19 PM

Quant Time: Jan 16 18:32:26 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 16 15:00:30 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	8469	5.00	ng	0.00
7) Naphthalene-d8	10.91	136	40031	5.00	ng	0.00
12) Acenaphthene-d10	14.78	164	26839	5.00	ng	0.00
18) Phenanthrene-d10	17.53	188	59875	5.00	ng	0.00
24) Chrysene-d12	21.72	240	61002	5.00	ng	0.00
31) Perylene-d12	24.06	264	67990	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.64	112	9670	7.51	ng	-0.01
5) Phenol-d6	7.31	99	12794	6.93	ng	-0.02
8) Nitrobenzene-d5	9.29	82	13969	6.23	ng	-0.05
13) 2,4,6-Tribromophenol	16.27	330	4012	6.61	ng	-0.02
14) 2-Fluorobiphenyl	13.40	172	32250	5.02	ng	-0.02
26) Terphenyl-d14	20.14	244	43932	5.87	ng	-0.02

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.37	88	5652	4.55	ng	92
3) n-Nitrosodimethylamine	3.72	42	7920	7.63	ng	92
6) bis(2-Chloroethyl)ether	7.54	93	12742	6.65	ng	# 83
9) Naphthalene	10.95	128	40236	4.90	ng	# 95
10) Hexachlorobutadiene	11.24	225	5888	4.72	ng	100
11) 2-Methylnaphthalene	12.59	142	26407m	5.05	ng	
15) Acenaphthylene	14.49	152	188561	5.09	ng	100
16) Acenaphthene	14.85	154	28332	4.64	ng	89
17) Fluorene	15.83	166	37419	4.99	ng	99
19) Hexachlorobenzene	16.87	284	9904m	3.87	ng	
21) Phenanthrene	17.58	178	61255	4.06	ng	# 94
22) Anthracene	17.67	178	61805	4.48	ng	99
23) Fluoranthene	19.58	202	296184	4.54	ng	99
25) Pyrene	19.93	202	314422	5.85	ng	99
27) Benzo(a)anthracene	21.70	228	306380	5.59	ng	# 71
28) Chrysene	21.75	228	314485m	4.95	ng	
29) Bis(2-ethylhexyl)phthalate	21.61	149	55967	9.44	ng	# 71
30) Indeno(1,2,3-cd)pyrene	26.53	276	390103	6.80	ng	97
32) Benzo(b)fluoranthene	23.34	252	87878	5.62	ng	# 97
33) Benzo(k)fluoranthene	23.38	252	81193	4.66	ng	95
34) Benzo(a)pyrene	23.95	252	81625	5.41	ng	# 94
35) Dibenzo(a,h)anthracene	26.55	278	80715	5.63	ng	95
36) Benzo(g,h,i)perylene	27.28	276	333268	5.43	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE011617\
Data File : BE092654.D
Acq On : 16 Jan 2017 16:28
Operator : UM/SJ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM

Quant Time: Jan 16 18:32:26 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 16 15:00:30 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. -0.00 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion:152 Resp: 8469

Ion Ratio Lower Upper

152 100


150 143.4 106.0 159.0

115 44.8 31.2 46.8

Manual Integrations
APPROVED

mohammad




1/17/2017 4:40:19 PM



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 4.55 ng

RT: 3.37 min Scan# 68

Delta R.T. 0.00 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28


Tgt Ion: 88 Resp: 5652

Ion Ratio Lower Upper

88 100

58 88.1 63.6 95.4



43 37.5 28.0 42.0




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 7.63 ng

RT: 3.72 min Scan# 98

Delta R.T. -0.02 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	42	Resp:	7920
Ion	Ratio	Lower	Upper
42	100		
74	99.2	86.0	129.0
44	7.2	5.1	7.7

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#4

2-Fluorophenol

Concen: 7.51 ng




RT: 5.64 min Scan# 358


Delta R.T. -0.01 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Tgt Ion:	112	Resp:	9670
Ion	Ratio	Lower	Upper
112	100		
64	68.7	53.5	80.3
63	36.8	27.9	41.9





#5

Phenol-d6

Concen: 6.93 ng

RT: 7.31 min Scan# 539

Delta R.T. -0.02 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	99	Resp:	12794
Ion	Ratio	Lower	Upper
99	100		
42	0.0	16.0	24.0#
71	0.0	30.6	45.8#

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#6

bis(2-Chloroethyl)ether

Concen: 6.65 ng




RT: 7.54 min Scan# 549

Delta R.T. -0.02 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Tgt Ion:	93	Resp:	12742
Ion	Ratio	Lower	Upper
93	100		
63	86.1	71.6	107.4
95	0.0	24.0	36.0#





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.91 min Scan# 706

Delta R.T. -0.00 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	136	Resp:	40031
Ion Ratio	Lower	Upper	
136	100		
137	11.2	8.2	12.4
54	7.3	9.6	14.4#
68	5.5	6.7	10.1#

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#8

Nitrobenzene-d5

Concen: 6.23 ng




RT: 9.29 min Scan# 625

Delta R.T. -0.05 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Tgt Ion:	82	Resp:	13969
Ion Ratio	Lower	Upper	
82	100		
128	13.2	39.0	58.4#
54	75.2	44.8	67.2#





#9

Naphthalene

Concen: 4.90 ng

RT: 10.95 min Scan# 708

Delta R.T. -0.02 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E





ClientSampleId :

SSTDICC005

Tot Ion:	128	Resp:	40236
Ion Ratio	Lower	Upper	
128	100		
129	10.2	11.2	16.8#
127	14.5	11.6	17.4

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#10

Hexachlorobutadiene

Concen: 4.72 ng




RT: 11.24 min Scan# 723


Delta R.T. -0.02 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Tgt Ion:	225	Resp:	5888
Ion Ratio	Lower	Upper	
225	100		
223	62.9	50.6	76.0
227	64.0	51.4	77.0









#11
2-Methylnaphthalene
Concen: 5.05 ng
RT: 12.59 min Scan# 811
Delta R.T. -0.02 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
142	100		
141	85.2	73.7	110.5
115	33.2	34.0	51.0#




Manual Integrations
APPROVED


mohammad
1/17/2017 4:40:19 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.78 min Scan# 978
Delta R.T. 0.00 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.3	71.4	107.0
160	46.1	27.4	41.2#









#13
 2,4,6-Tribromophenol
 Concen: 6.61 ng
 RT: 16.27 min Scan# 1085
 Delta R.T. -0.02 min
 Lab File: BE092654.D
 Acq: 16 Jan 2017 16:28

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion:330 Resp: 4012
 Ion Ratio Lower Upper
 330 100
 332 95.8 75.4 113.0
 141 44.3 31.0 46.6




Manual Integrations
 APPROVED

mohammad
 1/17/2017 4:40:19 PM



#14
 2-Fluorobiphenyl
 Concen: 5.02 ng
 RT: 13.40 min Scan# 881
 Delta R.T. -0.02 min
 Lab File: BE092654.D
 Acq: 16 Jan 2017 16:28

Tgt Ion:172 Resp: 32250
 Ion Ratio Lower Upper
 172 100
 171 34.8 30.1 45.1
 170 23.3 21.8 32.6









#15
Acenaphthylene
Concen: 5.09 ng
RT: 14.49 min Scan# 961
Delta R.T. -0.02 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.1	10.4	15.6




Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#16
Acenaphthene
Concen: 4.64 ng
RT: 14.85 min Scan# 982
Delta R.T. 0.00 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Tgt Ion	Ratio	Lower	Upper
154	100		
153	463.8	350.8	526.2
152	234.0	171.1	256.7





#17

Fluorene

Concen: 4.99 ng

RT: 15.83 min Scan# 1047

Delta R.T. -0.01 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion:166 Resp: 37419

Ion Ratio Lower Upper

166 100

165 98.9 78.2 117.4





167 13.4 11.0 16.4

Manual Integrations

APPROVED

mohammad

1/17/2017 4:40:19 PM



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.53 min Scan# 1152

Delta R.T. -0.00 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28




Tgt Ion:188 Resp: 59875

Ion Ratio Lower Upper

188 100

94 7.8 3.2 4.8#

80 7.4 2.6 3.8#









#19
Hexachlorobenzene
Concen: 3.87 ng m
RT: 16.87 min Scan# 1123
Delta R.T. -0.00 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	0.0	27.9	41.9#




Manual Integrations
APPROVED


mohammad
1/17/2017 4:40:19 PM



#21
Phenanthrene
Concen: 4.06 ng
RT: 17.58 min Scan# 1154
Delta R.T. 0.00 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.8	23.8
179	15.9	16.0	24.0#





#22

Anthracene

Concen: 4.48 ng

RT: 17.67 min Scan# 1158

Delta R.T. -0.02 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E





ClientSampleId :

SSTDICC005

Tot Ion:	178	Resp:	61805
Ion Ratio	Lower	Upper	
178	100		
176	18.2	15.0	22.4
179	15.8	12.2	18.2

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#23

Fluoranthene

Concen: 4.54 ng




RT: 19.58 min Scan# 1264


Delta R.T. -0.02 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Tgt Ion:	202	Resp:	296184
Ion Ratio	Lower	Upper	
202	100		
101	2.8	2.2	3.4
203	17.4	13.7	20.5









#24
Chrysene-d12
Concen: 5.00 ng
RT: 21.72 min Scan# 1374
Delta R.T. -0.00 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion:240 Resp: 61002
Ion Ratio Lower Upper
240 100
120 1.0 2.6 4.0#
236 20.2 24.6 37.0#




Manual Integrations
APPROVED


mohammad
1/17/2017 4:40:19 PM



#25
Pyrene
Concen: 5.85 ng
RT: 19.93 min Scan# 1285
Delta R.T. -0.02 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Tgt Ion:202 Resp: 314422
Ion Ratio Lower Upper
202 100
200 5.5 4.2 6.4
203 17.1 13.9 20.9









#26
 Terphenyl-d14
 Concen: 5.87 ng
 RT: 20.14 min Scan# 1297
 Delta R.T. -0.02 min
 Lab File: BE092654.D
 Acq: 16 Jan 2017 16:28

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion	Ratio	Lower	Upper
244	100		
212	10.3	6.7	10.1#
122	13.3	3.6	5.4#




Manual Integrations
 APPROVED


mohammad
 1/17/2017 4:40:19 PM



#27
 Benzo(a)anthracene
 Concen: 5.59 ng
 RT: 21.70 min Scan# 1373
 Delta R.T. 0.00 min
 Lab File: BE092654.D
 Acq: 16 Jan 2017 16:28

Tgt Ion	Ratio	Lower	Upper
228	100		
226	15.1	23.6	35.4#
229	30.4	13.3	19.9#





#28

Chrysene

Concen: 4.95 ng m

RT: 21.75 min Scan# 1375

Delta R.T. 0.00 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E





ClientSampleId :

SSTDICC005

Tot Ion:	228	Resp:	314485
Ion Ratio	Lower	Upper	
228	100		
226	22.8	36.3	54.5#
229	21.5	10.6	16.0#

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#29

Bis(2-ethylhexyl)phthalate

Concen: 9.44 ng




RT: 21.61 min Scan# 1369


Delta R.T. -0.00 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Tgt Ion:	149	Resp:	55967
Ion Ratio	Lower	Upper	
149	100		
167	27.1	16.0	24.0#
279	0.0	16.0	24.0#









#30
Indeno(1,2,3-cd)pyrene
Concen: 6.80 ng
RT: 26.53 min Scan# 1712
Delta R.T. -0.02 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
276	100		
138	28.0	20.3	30.5
277	24.8	19.7	29.5




Manual Integrations
APPROVED


mohammad
1/17/2017 4:40:19 PM



#31
Perylene-d12
Concen: 5.00 ng
RT: 24.06 min Scan# 1542
Delta R.T. 0.00 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	20.1	30.1
265	22.4	17.9	26.9





#32

Benzo(b)fluoranthene

Concen: 5.62 ng

RT: 23.34 min Scan# 1479

Delta R.T. -0.01 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Instrument :

BNA_E





ClientSampleId :

SSTDICC005

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.6	18.7	28.1
125	9.9	10.5	15.7

Manual Integrations
APPROVED

mohammad
1/17/2017 4:40:19 PM



#33

Benzo(k)fluoranthene

Concen: 4.66 ng




RT: 23.38 min Scan# 1483


Delta R.T. -0.01 min

Lab File: BE092654.D

Acq: 16 Jan 2017 16:28

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.6	20.3	30.5
125	10.0	9.6	14.4









#34
Benzo(a)pyrene
Concen: 5.41 ng
RT: 23.95 min Scan# 1532
Delta R.T. -0.01 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.7	19.0	28.6
125	11.6	11.8	17.8




Manual Integrations
APPROVED




mohammad
1/17/2017 4:40:19 PM



#35
Dibenzo(a,h)anthracene
Concen: 5.63 ng
RT: 26.55 min Scan# 1713
Delta R.T. -0.03 min
Lab File: BE092654.D
Acq: 16 Jan 2017 16:28

Tgt Ion	Ratio	Lower	Upper
278	100		
139	15.0	13.8	20.8
279	24.4	21.4	32.2





#36
 Benzo(a,h,i)perylene
 Concen: 5.43 ng
 RT: 27.28 min Scan# 1756
 Delta R.T. -0.03 min
 Lab File: BE092654.D
 Acq: 16 Jan 2017 16:28

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.1	19.8	29.8
138	24.1	19.8	29.6

Manual Integrations
 APPROVED

mohammad
 1/17/2017 4:40:19 PM

