

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011620\
 Data File : BE101070.D
 Acq On : 16 Jan 2020 10:37
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 17 07:52:03 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	1.020	1.044	-2.4	92	0.00
3	n-Nitrosodimethylamine	0.556	0.542	2.5	94	0.00
4 S	2-Fluorophenol	0.900	0.979	-8.8	91	0.00
5 S	Phenol-d6	1.139	1.154	-1.3	95	0.00
6	bis(2-Chloroethyl)ether	1.357	1.212	10.7	81	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.286	0.264	7.7	87	0.00
9	Naphthalene	4.441	4.565	-2.8	85	0.00
10	Hexachlorobutadiene	0.188	0.205	-9.0	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.750	1.7	83	0.00
12	2-Methylnaphthalene	0.672	0.619	7.9	79	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.113	0.105	7.1	85	0.00
15 S	2-Fluorobiphenyl	1.519	1.566	-3.1	81	-0.01
16	Acenaphthylene	1.629	1.679	-3.1	91	-0.01
17	Acenaphthene	1.166	1.190	-2.1	83	-0.01
18	Fluorene	1.474	1.403	4.8	77	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
20	4-Bromophenyl-phenylether	0.195	0.182	6.7	76	0.00
21	Hexachlorobenzene	0.214	0.222	-3.7	80	-0.01
22	Pentachlorophenol	0.053	0.037	30.2#	72	0.00
23	Phenanthrene	1.092	1.055	3.4	76	0.00
24	Anthracene	1.031	1.025	0.6	81	0.00
25 SURR	Fluoranthene-d10	5.555	5.422	2.4	78	0.00
26	Fluoranthene	1.368	1.331	2.7	81	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	87	-0.01
28	Pvrene	1.489	1.407	5.5	80	0.00
29 S	Terphenyl-d14	0.982	0.878	10.6	77	0.00
30	Benzo(a)anthracene	1.141	1.084	5.0	86	0.00
31	Chrysene	1.696	1.788	-5.4	89	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	2.027	-6.6	107	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.705	-27.2#	114	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	114	0.00
35	Benzo(b)fluoranthene	1.122	0.934	16.8	102	0.00
36	Benzo(k)fluoranthene	1.626	1.341	17.5	94	0.00
37 C	Benzo(a)pyrene	1.085	1.050	3.2	114	0.00
38	Dibenzo(a,h)anthracene	0.998	0.984	1.4	122	-0.02
39	Benzo(a,h,i)perylene	1.227	1.165	5.1	106	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011620\
Data File : BE101070.D
Acq On : 16 Jan 2020 10:37
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 17 07:52:03 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			