

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011620\
 Data File : BE101070.D
 Acq On : 16 Jan 2020 10:37
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 17 07:52:03 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	0.00
2	1,4-Dioxane	0.400	0.447	-11.7	92	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	94	0.00
4 S	2-Fluorophenol	0.400	0.435	-8.7	91	0.00
5 S	Phenol-d6	0.400	0.405	-1.3	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	81	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.369	7.8	87	0.00
9	Naphthalene	0.400	0.411	-2.7	85	0.00
10	Hexachlorobutadiene	0.400	0.436	-9.0	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	83	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	79	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.400	0.371	7.3	85	0.00
15 S	2-Fluorobiphenyl	0.400	0.412	-3.0	81	-0.01
16	Acenaphthylene	0.400	0.412	-3.0	91	-0.01
17	Acenaphthene	0.400	0.408	-2.0	83	-0.01
18	Fluorene	0.400	0.381	4.8	77	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	78	0.00
20	4-Bromophenyl-phenylether	0.400	0.374	6.5	76	0.00
21	Hexachlorobenzene	0.400	0.414	-3.5	80	-0.01
22	Pentachlorophenol	0.400	0.449	-12.2	72	0.00
23	Phenanthrene	0.400	0.386	3.5	76	0.00
24	Anthracene	0.400	0.398	0.5	81	0.00
25 SURR	Fluoranthene-d10	0.400	0.390	2.5	78	0.00
26	Fluoranthene	0.400	0.389	2.8	81	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	87	-0.01
28	Pvrene	0.400	0.378	5.5	80	0.00
29 S	Terphenyl-d14	0.400	0.358	10.5	77	0.00
30	Benzo(a)anthracene	0.400	0.380	5.0	86	0.00
31	Chrysene	0.400	0.422	-5.5	89	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.426	-6.5	107	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.509	-27.2#	114	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	114	0.00
35	Benzo(b)fluoranthene	0.400	0.333	16.8	102	0.00
36	Benzo(k)fluoranthene	0.400	0.330	17.5	94	0.00
37 C	Benzo(a)pyrene	0.400	0.387	3.3	114	0.00
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	122	-0.02
39	Benzo(a,h,i)perylene	0.400	0.380	5.0	106	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011620\
Data File : BE101070.D
Acq On : 16 Jan 2020 10:37
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 17 07:52:03 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				