

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011620\
 Data File : BE101084.D
 Acq On : 16 Jan 2020 20:54
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 17 10:44:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	-0.01
2	1,4-Dioxane	1.020	0.888	12.9	95	0.00
3	n-Nitrosodimethylamine	0.556	0.553	0.5	117	-0.01
4 S	2-Fluorophenol	0.900	1.104	-22.7	125	-0.01
5 S	Phenol-d6	1.139	1.350	-18.5	134	-0.01
6	bis(2-Chloroethyl)ether	1.357	1.352	0.4	110	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.286	0.289	-1.0	117	-0.01
9	Naphthalene	4.441	4.659	-4.9	106	-0.01
10	Hexachlorobutadiene	0.188	0.205	-9.0	109	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.960	-25.8#	130	-0.02
12	2-Methylnaphthalene	0.672	0.719	-7.0	113	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.113	0.104	8.0	117	-0.01
15 S	2-Fluorobiphenyl	1.519	1.464	3.6	104	-0.02
16	Acenaphthylene	1.629	1.738	-6.7	129	-0.02
17	Acenaphthene	1.166	1.218	-4.5	117	-0.02
18	Fluorene	1.474	1.521	-3.2	115	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.195	0.206	-5.6	124	-0.01
21	Hexachlorobenzene	0.214	0.211	1.4	110	-0.01
22	Pentachlorophenol	0.053	0.014	73.6#	39#	0.00
23	Phenanthrene	1.092	1.073	1.7	112	0.00
24	Anthracene	1.031	1.106	-7.3	127	0.00
25 SURR	Fluoranthene-d10	5.555	7.648	-37.7#	160#	-0.01
26	Fluoranthene	1.368	1.423	-4.0	124	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	128	-0.01
28	Pvrene	1.489	1.495	-0.4	125	0.00
29 S	Terphenyl-d14	0.982	0.935	4.8	120	-0.01
30	Benzo(a)anthracene	1.141	1.346	-18.0	157#	0.00
31	Chrysene	1.696	1.722	-1.5	126	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	3.648	-91.8#	281#	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.789	-33.5#	175#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	150#	-0.01
35	Benzo(b)fluoranthene	1.122	1.143	-1.9	164#	-0.01
36	Benzo(k)fluoranthene	1.626	1.593	2.0	147	0.00
37 C	Benzo(a)pyrene	1.085	1.200	-10.6	172#	0.00
38	Dibenzo(a,h)anthracene	0.998	1.124	-12.6	183#	-0.01
39	Benzo(a,h,i)perylene	1.227	1.205	1.8	144	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011620\
Data File : BE101084.D
Acq On : 16 Jan 2020 20:54
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 17 10:44:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			