

Data Path : Z:\HPCHEM1\BNA E\DATA\BE011917\
 Data File : BE092658.D
 Acq On : 19 Jan 2017 13:44
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 19 14:51:53 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.775	0.852	-9.9	122	0.01
3	n-Nitrosodimethylamine	0.902	0.790	12.4	117	0.02
4 S	2-Fluorophenol	1.116	0.948	15.1	116	0.02
5 S	Phenol-d6	1.368	1.152	15.8	119	0.00
6	bis(2-Chloroethyl)ether	1.453	1.249	14.0	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.325	0.281	13.5	122	0.00
9	Naphthalene	1.044	1.036	0.8	126	0.00
10	Hexachlorobutadiene	0.154	0.158	-2.6	126	0.00
11	2-Methylnaphthalene	0.648	0.619	4.5	125	0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
13 S	2,4,6-Tribromophenol	0.126	0.119	5.6	137	0.00
14 S	2-Fluorobiphenyl	1.230	1.262	-2.6	128	0.00
15	Acenaphthylene	6.890	6.813	1.1	129	0.00
16	Acenaphthene	1.105	1.149	-4.0	127	0.00
17	Fluorene	1.387	1.357	2.2	126	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
19	Hexachlorobenzene	0.189	0.210	-11.1	123	0.00
20	Pentachlorophenol	0.051	0.042	17.6	103	0.00
21	Phenanthrene	1.104	1.178	-6.7	127	0.00
22	Anthracene	1.018	1.102	-8.3	140	0.00
23	Fluoranthene	4.895	4.909	-0.3	127	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
25	Pyrene	4.929	4.524	8.2	126	0.00
26 S	Terphenyl-d14	0.707	0.662	6.4	125	0.00
27	Benzo(a)anthracene	4.926	4.572	7.2	123	0.00
28	Chrysene	5.542	6.022	-8.7	134	0.00
29	Bis(2-ethylhexyl)phthalate	0.913	0.571	37.5#	100	0.00
30	Indeno(1,2,3-cd)pyrene	5.987	5.452	8.9	119	0.00
31 I	Perylene-d12	1.000	1.000	0.0	112	0.00
32	Benzo(b)fluoranthene	1.210	1.171	3.2	118	0.00
33	Benzo(k)fluoranthene	1.215	1.198	1.4	129	0.00
34 C	Benzo(a)pyrene	1.136	1.075	5.4	123	0.00
35	Dibenzo(a,h)anthracene	1.112	1.044	6.1	119	0.00
36	Benzo(g,h,i)perylene	4.699	4.485	4.6	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0