

Data Path : Z:\HPCHEM1\BNA E\Data\BE011917\
 Data File : BE092669.D
 Acq On : 19 Jan 2017 22:36
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 20 02:24:18 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.775	0.877	-13.2	127	0.01
3	n-Nitrosodimethylamine	0.902	0.647	28.3#	97	0.02
4 S	2-Fluorophenol	1.116	0.955	14.4	119	0.02
5 S	Phenol-d6	1.368	1.180	13.7	124	0.00
6	bis(2-Chloroethyl)ether	1.453	1.203	17.2	117	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.325	0.270	16.9	125	0.00
9	Naphthalene	1.044	1.049	-0.5	135	0.00
10	Hexachlorobutadiene	0.154	0.155	-0.6	132	-0.02
11	2-Methylnaphthalene	0.648	0.640	1.2	137	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
13 S	2,4,6-Tribromophenol	0.126	0.122	3.2	159#	0.01
14 S	2-Fluorobiphenyl	1.230	1.204	2.1	138	0.00
15	Acenaphthylene	6.890	6.849	0.6	147	0.00
16	Acenaphthene	1.105	1.146	-3.7	143	0.00
17	Fluorene	1.387	1.381	0.4	145	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	141	0.00
19	Hexachlorobenzene	0.189	0.189	0.0	135	0.00
20	Pentachlorophenol	0.051	0.035	31.4#	104	0.14
21	Phenanthrene	1.104	1.157	-4.8	152#	0.00
22	Anthracene	1.018	1.014	0.4	157#	-0.02
23	Fluoranthene	4.895	4.945	-1.0	156#	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
25	Pyrene	4.929	5.112	-3.7	155#	0.00
26 S	Terphenyl-d14	0.707	0.713	-0.8	147	0.00
27	Benzo(a)anthracene	4.926	5.049	-2.5	148	0.00
28	Chrysene	5.542	5.749	-3.7	139	0.00
29	Bis(2-ethylhexyl)phthalate	0.913	0.696	23.8	133	0.00
30	Indeno(1,2,3-cd)pyrene	5.987	5.530	7.6	132	0.00
31 I	Perylene-d12	1.000	1.000	0.0	123	0.00
32	Benzo(b)fluoranthene	1.210	1.139	5.9	127	-0.01
33	Benzo(k)fluoranthene	1.215	1.310	-7.8	155#	-0.01
34 C	Benzo(a)pyrene	1.136	1.122	1.2	141	-0.01
35	Dibenzo(a,h)anthracene	1.112	1.043	6.2	131	-0.02
36	Benzo(g,h,i)perylene	4.699	4.441	5.5	129	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0