

Data Path : Z:\HPCHEM1\BNA E\Data\BE011917\
 Data File : BE092669.D
 Acq On : 19 Jan 2017 22:36
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 20 02:24:18 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	118	0.00
2	1,4-Dioxane	0.400	0.459	-14.8	127	0.01
3	n-Nitrosodimethylamine	0.400	0.287	28.3#	97	0.02
4 S	2-Fluorophenol	0.400	0.342	14.5	119	0.02
5 S	Phenol-d6	0.400	0.345	13.8	124	0.00
6	bis(2-Chloroethyl)ether	0.400	0.331	17.3	117	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.400	0.333	16.8	125	0.00
9	Naphthalene	0.400	0.402	-0.5	135	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	132	-0.02
11	2-Methylnaphthalene	0.400	0.395	1.3	137	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	131	0.00
13 S	2,4,6-Tribromophenol	0.400	0.457	-14.2	159	0.01
14 S	2-Fluorobiphenyl	0.400	0.392	2.0	138	0.00
15	Acenaphthylene	0.400	0.398	0.5	147	0.00
16	Acenaphthene	0.400	0.415	-3.7	143	0.00
17	Fluorene	0.400	0.398	0.5	145	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	141	0.00
19	Hexachlorobenzene	0.400	0.378	5.5	135	0.00
20	Pentachlorophenol	0.400	0.271	32.3#	104	0.14
21	Phenanthrene	0.400	0.419	-4.7	152	0.00
22	Anthracene	0.400	0.398	0.5	157	-0.02
23	Fluoranthene	0.400	0.404	-1.0	156	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	125	0.00
25	Pyrene	0.400	0.415	-3.7	155	0.00
26 S	Terphenyl-d14	0.400	0.404	-1.0	147	0.00
27	Benzo(a)anthracene	0.400	0.410	-2.5	148	0.00
28	Chrysene	0.400	0.415	-3.7	139	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.481	-20.2	133	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	132	0.00
31 I	Perylene-d12	5.000	5.000	0.0	123	0.00
32	Benzo(b)fluoranthene	0.400	0.377	5.8	127	-0.01
33	Benzo(k)fluoranthene	0.400	0.431	-7.7	155	-0.01
34 C	Benzo(a)pyrene	0.400	0.395	1.3	141	-0.01
35	Dibenzo(a,h)anthracene	0.400	0.375	6.3	131	-0.02
36	Benzo(g,h,i)perylene	0.400	0.378	5.5	129	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0