

Data Path : U:\HPCHEM1\BNA E\DATA\BE012018\
 Data File : BE095085.D
 Acq On : 20 Jan 2018 12:23
 Operator : SJ/JU
 Sample : J1216-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 B0AE6

Manual Integrations
 APPROVED

Sohil
 1/22/2018 5:22:29 PM

Quant Time: Jan 22 12:14:51 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE011918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 22 04:29:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.46	152	4296	0.40	ng/ul	0.00
2) Naphthalene-d8	11.28	136	14481	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.04	164	9285	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.73	188	17979m	0.40	ng/ul	0.00
16) Chrysene-d12	21.83	240	18979	0.40	ng/ul	0.00
20) Perylene-d12	24.48	264	15531m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.84	152	8406	0.35	ng/ul	0.00
14) Fluoranthene-d10	19.72	212	21964	0.39	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.33	128	5089	0.133	ng/ul#	91
5) 2-Methylnaphthalene	12.91	142	3112	0.116	ng/ul	98
7) Acenaphthylene	14.77	152	14481	0.348	ng/ul	97
8) Acenaphthene	15.10	153	8983	0.278	ng/ul	93
9) Fluorene	16.05	166	12527	0.363	ng/ul#	78
12) Phenanthrene	17.78	178	230057	4.819	ng/ul#	87
13) Anthracene	17.87	178	25120	0.603	ng/ul#	91
15) Fluoranthene	19.75	202	324344	5.316	ng/ul	84
17) Pyrene	20.10	202	326595	4.734	ng/ul#	80
18) Benzo(a)anthracene	21.81	228	132000	2.514	ng/ul#	86
19) Chrysene	21.87	228	123637	2.110	ng/ul	97
21) Benzo(b)fluoranthene	23.67	252	123606m	2.364	ng/ul	
22) Benzo(k)fluoranthene	23.71	252	43272m	0.897	ng/ul	
23) Benzo(a)pyrene	24.36	252	97350m	2.181	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.31	276	56378m	1.422	ng/ul	
25) Dibenzo(a,h)anthracene	27.32	278	14930m	0.527	ng/ul	
26) Benzo(a,h,i)perylene	28.19	276	48518m	1.231	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA E\DATA\BE012018\
Data File : BE095085.D
Acq On : 20 Jan 2018 12:23
Operator : SJ/JU
Sample : J1216-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
B0AE6

Manual Integrations
APPROVED

Sohil
1/22/2018 5:22:29 PM

Quant Time: Jan 22 12:14:51 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE011918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 22 04:29:25 2018
Response via : Initial Calibration

