

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE012120\
 Data File : BE101104.D
 Acq On : 21 Jan 2020 9:36
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 22 07:55:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	-0.01
2	1,4-Dioxane	1.020	0.876	14.1	64	0.00
3	n-Nitrosodimethylamine	0.556	0.466	16.2	68	-0.01
4 S	2-Fluorophenol	0.900	0.926	-2.9	72	-0.01
5 S	Phenol-d6	1.139	0.990	13.1	68	-0.01
6	bis(2-Chloroethyl)ether	1.357	1.055	22.3	59	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	71	0.00
8 S	Nitrobenzene-d5	0.286	0.249	12.9	69	0.00
9	Naphthalene	4.441	4.342	2.2	68	0.00
10	Hexachlorobutadiene	0.188	0.201	-6.9	73	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.773	-1.3	71	0.00
12	2-Methylnaphthalene	0.672	0.617	8.2	66	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
14 S	2,4,6-Tribromophenol	0.113	0.083	26.5#	70	0.00
15 S	2-Fluorobiphenyl	1.519	1.335	12.1	72	0.00
16	Acenaphthylene	1.629	1.471	9.7	83	0.00
17	Acenaphthene	1.166	1.071	8.1	78	-0.01
18	Fluorene	1.474	1.417	3.9	81	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4-Bromophenyl-phenylether	0.195	0.185	5.1	91	0.00
21	Hexachlorobenzene	0.214	0.191	10.7	81	-0.01
22	Pentachlorophenol	0.053	0.017	67.9#	38#	0.00
23	Phenanthrene	1.092	0.991	9.2	84	0.00
24	Anthracene	1.031	0.979	5.0	92	0.00
25 SURR	Fluoranthene-d10	5.555	6.003	-8.1	103	0.00
26	Fluoranthene	1.368	1.455	-6.4	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	-0.01
28	Pvrene	1.489	1.386	6.9	104	0.00
29 S	Terphenyl-d14	0.982	0.929	5.4	107	0.00
30	Benzo(a)anthracene	1.141	0.983	13.8	103	0.00
31	Chrysene	1.696	1.701	-0.3	112	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	1.379	27.5#	96	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.147	14.4	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	-0.01
35	Benzo(b)fluoranthene	1.122	0.924	17.6	91	0.00
36	Benzo(k)fluoranthene	1.626	1.361	16.3	86	0.00
37 C	Benzo(a)pyrene	1.085	1.023	5.7	100	0.00
38	Dibenzo(a,h)anthracene	0.998	0.947	5.1	106	-0.02
39	Benzo(a,h,i)perylene	1.227	1.127	8.1	92	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE012120\
Data File : BE101104.D
Acq On : 21 Jan 2020 9:36
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 22 07:55:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			