

Data Path : U:\HPCHEM1\BNA E\DATA\BE012318\
 Data File : BE095105.D
 Acq On : 23 Jan 2018 19:49
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 24 01:11:37 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE012318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 15:47:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.393	0.388	1.3	109	0.00
3	n-Nitrosodimethylamine	0.469	0.466	0.6	109	0.00
4 S	2-Fluorophenol	0.648	0.687	-6.0	117	0.00
5 S	Phenol-d6	0.982	1.052	-7.1	120	0.00
6	bis(2-Chloroethyl)ether	0.914	0.914	0.0	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.364	0.360	1.1	116	0.00
9	Naphthalene	4.577	4.544	0.7	110	0.00
10	Hexachlorobutadiene	0.205	0.238	-16.1	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.664	0.666	-0.3	112	0.00
12	2-Methylnaphthalene	0.785	0.784	0.1	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.097	0.123	-26.8#	147	0.00
15 S	2-Fluorobiphenyl	1.758	1.704	3.1	111	0.00
16	Acenaphthylene	2.141	2.111	1.4	117	0.00
17	Acenaphthene	1.393	1.381	0.9	115	0.00
18	Fluorene	1.740	1.688	3.0	113	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.235	0.231	1.7	113	0.00
21	Hexachlorobenzene	0.236	0.224	5.1	110	0.00
22	Pentachlorophenol	0.055	0.069	-25.5#	177#	0.00
23	Phenanthrene	1.236	1.208	2.3	112	0.00
24	Anthracene	1.102	1.077	2.3	115	0.00
25 SURR	Fluoranthene-d10	5.031	4.870	3.2	113	0.00
26	Fluoranthene	1.484	1.438	3.1	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
28	Pvrene	1.691	1.681	0.6	110	0.00
29 S	Terphenyl-d14	0.819	0.819	0.0	111	0.00
30	Benzo(a)anthracene	1.251	1.232	1.5	107	-0.01
31	Chrysene	1.265	1.258	0.6	118	0.00
32	Bis(2-ethylhexyl)phthalate	1.740	2.271	-30.5#	152#	0.00
33	Indeno(1,2,3-cd)pyrene	1.150	1.217	-5.8	116	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.419	1.383	2.5	112	0.00
36	Benzo(k)fluoranthene	1.452	1.377	5.2	111	0.00
37 C	Benzo(a)pyrene	1.301	1.291	0.8	114	0.00
38	Dibenzo(a,h)anthracene	1.095	1.119	-2.2	117	0.00
39	Benzo(a,h,i)perylene	1.187	1.143	3.7	113	0.00

Data Path : U:\HPCHEM1\BNA E\DATA\BE012318\
Data File : BE095105.D
Acq On : 23 Jan 2018 19:49
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 24 01:11:37 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE012318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 15:47:25 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			