

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012417\
 Data File : BE092671.D
 Acq On : 24 Jan 2017 18:31
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 25 13:12:36 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.775	0.676	12.8	93	0.00
3	n-Nitrosodimethylamine	0.902	0.518	42.6#	73	0.05
4 S	2-Fluorophenol	1.116	0.795	28.8#	94	0.02
5 S	Phenol-d6	1.368	1.051	23.2	104	0.02
6	bis(2-Chloroethyl)ether	1.453	1.243	14.5	114	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.325	0.265	18.5	111	0.00
9	Naphthalene	1.044	1.016	2.7	119	0.00
10	Hexachlorobutadiene	0.154	0.151	1.9	117	0.00
11	2-Methylnaphthalene	0.648	0.599	7.6	117	0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
13 S	2,4,6-Tribromophenol	0.126	0.114	9.5	135	0.03
14 S	2-Fluorobiphenyl	1.230	1.153	6.3	121	0.00
15	Acenaphthylene	6.890	6.454	6.3	126	0.00
16	Acenaphthene	1.105	1.105	0.0	126	0.00
17	Fluorene	1.387	1.311	5.5	126	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
19	Hexachlorobenzene	0.189	0.201	-6.3	122	0.00
20	Pentachlorophenol	0.051	0.001	98.0#	3#	0.00
21	Phenanthrene	1.104	1.153	-4.4	129	0.00
22	Anthracene	1.018	0.978	3.9	129	0.00
23	Fluoranthene	4.895	4.934	-0.8	133	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
25	Pyrene	4.929	4.467	9.4	134	0.00
26 S	Terphenyl-d14	0.707	0.630	10.9	128	0.00
27	Benzo(a)anthracene	4.926	4.595	6.7	134	0.00
28	Chrysene	5.542	5.778	-4.3	138	0.00
29	Bis(2-ethylhexyl)phthalate	0.913	0.485	46.9#	92	0.00
30	Indeno(1,2,3-cd)pyrene	5.987	5.356	10.5	126	0.00
31 I	Perylene-d12	1.000	1.000	0.0	121	0.00
32	Benzo(b)fluoranthene	1.210	1.171	3.2	127	0.00
33	Benzo(k)fluoranthene	1.215	1.148	5.5	133	0.00
34 C	Benzo(a)pyrene	1.136	1.051	7.5	129	0.00
35	Dibenzo(a,h)anthracene	1.112	1.017	8.5	125	0.00
36	Benzo(g,h,i)perylene	4.699	4.400	6.4	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0