

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012417\
 Data File : BE092682.D
 Acq On : 25 Jan 2017 1:26
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 25 13:48:21 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	147	0.00
2	1,4-Dioxane	0.775	0.686	11.5	124	0.01
3	n-Nitrosodimethylamine	0.902	0.571	36.7#	107	0.03
4 S	2-Fluorophenol	1.116	0.965	13.5	150	0.00
5 S	Phenol-d6	1.368	1.337	2.3	175#	0.00
6	bis(2-Chloroethyl)ether	1.453	1.216	16.3	147	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	165#	0.00
8 S	Nitrobenzene-d5	0.325	0.279	14.2	171#	-0.02
9	Naphthalene	1.044	1.015	2.8	174#	-0.02
10	Hexachlorobutadiene	0.154	0.153	0.6	173#	-0.02
11	2-Methylnaphthalene	0.648	0.637	1.7	182#	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	174#	0.00
13 S	2,4,6-Tribromophenol	0.126	0.110	12.7	191#	0.02
14 S	2-Fluorobiphenyl	1.230	1.159	5.8	177#	-0.01
15	Acenaphthylene	6.890	6.913	-0.3	197#	-0.02
16	Acenaphthene	1.105	1.111	-0.5	184#	0.00
17	Fluorene	1.387	1.324	4.5	185#	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	172#	0.00
19	Hexachlorobenzene	0.189	0.181	4.2	159#	0.00
20	Pentachlorophenol	0.051	0.032	37.3#	117	0.14
21	Phenanthrene	1.104	1.108	-0.4	179#	0.00
22	Anthracene	1.018	1.054	-3.5	200#	-0.02
23	Fluoranthene	4.895	4.572	6.6	177#	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	146	0.00
25	Pyrene	4.929	4.998	-1.4	177#	-0.02
26 S	Terphenyl-d14	0.707	0.710	-0.4	170#	0.00
27	Benzo(a)anthracene	4.926	4.886	0.8	167#	0.00
28	Chrysene	5.542	5.446	1.7	154#	0.00
29	Bis(2-ethylhexyl)phthalate	0.913	0.904	1.0	201#	0.00
30	Indeno(1,2,3-cd)pyrene	5.987	5.418	9.5	151#	0.00
31 I	Perylene-d12	1.000	1.000	0.0	135	0.00
32	Benzo(b)fluoranthene	1.210	1.181	2.4	143	-0.01
33	Benzo(k)fluoranthene	1.215	1.283	-5.6	166#	-0.01
34 C	Benzo(a)pyrene	1.136	1.153	-1.5	158#	0.00
35	Dibenzo(a,h)anthracene	1.112	1.105	0.6	151#	-0.02
36	Benzo(g,h,i)perylene	4.699	4.590	2.3	146	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0