

Data Path : Z:\HPCHEM1\BNA E\Data\BE012517\
 Data File : BE092684.D
 Acq On : 25 Jan 2017 14:19
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 25 17:25:23 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	154	0.00
2	1,4-Dioxane	0.400	0.341	14.8	129	0.00
3	n-Nitrosodimethylamine	0.400	0.257	35.8#	113	0.02
4 S	2-Fluorophenol	0.400	0.334	16.5	151	0.00
5 S	Phenol-d6	0.400	0.408	-2.0	191	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.356	11.0	163	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	179	0.00
8 S	Nitrobenzene-d5	0.400	0.339	15.3	183	-0.02
9	Naphthalene	0.400	0.390	2.5	189	-0.02
10	Hexachlorobutadiene	0.400	0.392	2.0	185	-0.02
11	2-Methylnaphthalene	0.400	0.396	1.0	198	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	187	0.00
13 S	2,4,6-Tribromophenol	0.400	0.422	-5.5	203	0.04
14 S	2-Fluorobiphenyl	0.400	0.382	4.5	192	-0.01
15	Acenaphthylene	0.400	0.398	0.5	210	-0.02
16	Acenaphthene	0.400	0.403	-0.8	198	0.00
17	Fluorene	0.400	0.385	3.8	200	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	182	0.00
19	Hexachlorobenzene	0.400	0.360	10.0	168	0.00
20	Pentachlorophenol	0.400	0.001	99.8#	0	0.02
21	Phenanthrene	0.400	0.412	-3.0	194	0.00
22	Anthracene	0.400	0.409	-2.2	209	-0.02
23	Fluoranthene	0.400	0.388	3.0	195	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	154	0.00
25	Pyrene	0.400	0.415	-3.7	191	-0.02
26 S	Terphenyl-d14	0.400	0.401	-0.3	180	0.00
27	Benzo(a)anthracene	0.400	0.370	7.5	165	0.00
28	Chrysene	0.400	0.390	2.5	161	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.537	-34.3#	197	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.366	8.5	161	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	142	0.00
32	Benzo(b)fluoranthene	0.400	0.383	4.3	148	-0.01
33	Benzo(k)fluoranthene	0.400	0.431	-7.7	178	-0.01
34 C	Benzo(a)pyrene	0.400	0.414	-3.5	170	0.00
35	Dibenzo(a,h)anthracene	0.400	0.398	0.5	159	-0.02
36	Benzo(g,h,i)perylene	0.400	0.390	2.5	153	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0