

Data Path : Z:\HPCHEM1\BNA E\Data\BE012517\
 Data File : BE092687.D
 Acq On : 25 Jan 2017 17:48
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 25 18:31:28 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	172	0.00
2	1,4-Dioxane	0.400	0.336	16.0	142	0.00
3	n-Nitrosodimethylamine	0.400	0.297	25.8#	146	0.01
4 S	2-Fluorophenol	0.400	0.383	4.3	194	0.00
5 S	Phenol-d6	0.400	0.437	-9.2	229	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.358	10.5	183	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	194	0.00
8 S	Nitrobenzene-d5	0.400	0.352	12.0	205	-0.02
9	Naphthalene	0.400	0.395	1.3	207	-0.02
10	Hexachlorobutadiene	0.400	0.398	0.5	203	-0.02
11	2-Methylnaphthalene	0.400	0.395	1.3	214	-0.02
12 I	Acenaphthene-d10	5.000	5.000	0.0	194	0.00
13 S	2,4,6-Tribromophenol	0.400	0.447	-11.7	229	0.01
14 S	2-Fluorobiphenyl	0.400	0.392	2.0	204	-0.02
15	Acenaphthylene	0.400	0.406	-1.5	221	-0.02
16	Acenaphthene	0.400	0.406	-1.5	207	0.00
17	Fluorene	0.400	0.392	2.0	212	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	190	0.00
19	Hexachlorobenzene	0.400	0.349	12.8	171	0.00
20	Pentachlorophenol	0.400	0.317	20.8	165	0.05
21	Phenanthrene	0.400	0.403	-0.8	198	0.00
22	Anthracene	0.400	0.427	-6.7	228	-0.02
23	Fluoranthene	0.400	0.406	-1.5	213	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	154	0.00
25	Pyrene	0.400	0.457	-14.2	211	-0.02
26 S	Terphenyl-d14	0.400	0.456	-14.0	205	0.00
27	Benzo(a)anthracene	0.400	0.399	0.3	178	0.00
28	Chrysene	0.400	0.421	-5.2	175	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.689	-72.2#	288	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.394	1.5	174	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	149	0.00
32	Benzo(b)fluoranthene	0.400	0.402	-0.5	163	-0.01
33	Benzo(k)fluoranthene	0.400	0.429	-7.2	186	-0.01
34 C	Benzo(a)pyrene	0.400	0.424	-6.0	183	0.00
35	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	173	-0.02
36	Benzo(g,h,i)perylene	0.400	0.409	-2.2	169	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0