

Data Path : U:\HPCHEM1\BNA E\Data\BE012518\
 Data File : BE095107.D
 Acq On : 25 Jan 2018 9:22
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 25 19:19:29 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE012318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 15:47:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.393	0.352	10.4	100	0.00
3	n-Nitrosodimethylamine	0.469	0.461	1.7	108	0.00
4 S	2-Fluorophenol	0.648	0.643	0.8	110	0.00
5 S	Phenol-d6	0.982	0.952	3.1	110	0.00
6	bis(2-Chloroethyl)ether	0.914	0.918	-0.4	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.364	0.349	4.1	110	0.00
9	Naphthalene	4.577	4.572	0.1	108	0.02
10	Hexachlorobutadiene	0.205	0.205	0.0	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.664	0.658	0.9	108	0.00
12	2-Methylnaphthalene	0.785	0.773	1.5	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.097	0.102	-5.2	116	0.00
15 S	2-Fluorobiphenyl	1.758	1.749	0.5	109	0.00
16	Acenaphthylene	2.141	2.074	3.1	109	0.00
17	Acenaphthene	1.393	1.380	0.9	109	0.00
18	Fluorene	1.740	1.683	3.3	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
20	4-Bromophenyl-phenylether	0.235	0.227	3.4	106	0.00
21	Hexachlorobenzene	0.236	0.234	0.8	109	0.00
22	Pentachlorophenol	0.055	0.064	-16.4	156#	0.00
23	Phenanthrene	1.236	1.208	2.3	107	0.00
24	Anthracene	1.102	1.066	3.3	108	0.00
25 SURR	Fluoranthene-d10	5.031	4.954	1.5	110	0.00
26	Fluoranthene	1.484	1.474	0.7	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
28	Pvrene	1.691	1.625	3.9	108	0.00
29 S	Terphenyl-d14	0.819	0.802	2.1	111	0.00
30	Benzo(a)anthracene	1.251	1.192	4.7	105	0.00
31	Chrysene	1.265	1.255	0.8	121	0.00
32	Bis(2-ethylhexyl)phthalate	1.740	1.773	-1.9	121	0.00
33	Indeno(1,2,3-cd)pyrene	1.150	1.146	0.3	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.419	1.407	0.8	115	0.00
36	Benzo(k)fluoranthene	1.452	1.347	7.2	109	0.00
37 C	Benzo(a)pyrene	1.301	1.280	1.6	113	0.00
38	Dibenzo(a,h)anthracene	1.095	1.050	4.1	111	0.00
39	Benzo(a,h,i)perylene	1.187	1.142	3.8	114	0.00

Data Path : U:\HPCHEM1\BNA E\Data\BE012518\
Data File : BE095107.D
Acq On : 25 Jan 2018 9:22
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 25 19:19:29 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE012318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 15:47:25 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			