

Data Path : U:\HPCHEM1\BNA E\Data\BE012518\
 Data File : BE095107.D
 Acq On : 25 Jan 2018 9:22
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 25 19:19:29 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE012318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 15:47:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.358	10.5	100	0.00
3	n-Nitrosodimethylamine	0.400	0.393	1.8	108	0.00
4 S	2-Fluorophenol	0.400	0.397	0.8	110	0.00
5 S	Phenol-d6	0.400	0.388	3.0	110	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.384	4.0	110	0.00
9	Naphthalene	0.400	0.400	0.0	108	0.02
10	Hexachlorobutadiene	0.400	0.400	0.0	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.396	1.0	108	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	108	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.400	0.397	0.8	116	0.00
15 S	2-Fluorobiphenyl	0.400	0.398	0.5	109	0.00
16	Acenaphthylene	0.400	0.387	3.3	109	0.00
17	Acenaphthene	0.400	0.396	1.0	109	0.00
18	Fluorene	0.400	0.387	3.3	107	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	108	0.00
20	4-Bromophenyl-phenylether	0.400	0.386	3.5	106	0.00
21	Hexachlorobenzene	0.400	0.396	1.0	109	0.00
22	Pentachlorophenol	0.400	0.526	-31.5#	156	0.00
23	Phenanthrene	0.400	0.391	2.3	107	0.00
24	Anthracene	0.400	0.387	3.3	108	0.00
25 SURR	Fluoranthene-d10	0.400	0.394	1.5	110	0.00
26	Fluoranthene	0.400	0.397	0.8	110	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	115	0.00
28	Pvrene	0.400	0.384	4.0	108	0.00
29 S	Terphenyl-d14	0.400	0.392	2.0	111	0.00
30	Benzo(a)anthracene	0.400	0.381	4.8	105	0.00
31	Chrysene	0.400	0.397	0.8	121	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.408	-2.0	121	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	111	0.00
34 I	Perylene-d12	0.400	0.400	0.0	115	0.00
35	Benzo(b)fluoranthene	0.400	0.397	0.8	115	0.00
36	Benzo(k)fluoranthene	0.400	0.371	7.3	109	0.00
37 C	Benzo(a)pyrene	0.400	0.394	1.5	113	0.00
38	Dibenzo(a,h)anthracene	0.400	0.384	4.0	111	0.00
39	Benzo(a,h,i)perylene	0.400	0.385	3.8	114	0.00

Data Path : U:\HPCHEM1\BNA E\Data\BE012518\
Data File : BE095107.D
Acq On : 25 Jan 2018 9:22
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 25 19:19:29 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE012318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 15:47:25 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				