

Data Path : U:\HPCHEM1\BNA E\DATA\BE012618\
 Data File : BE095149.D
 Acq On : 27 Jan 2018 3:52
 Operator : SJ/JU
 Sample : J1243-01
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6B25

Manual Integrations
 APPROVED

Sohil
 1/29/2018 6:31:02 PM

Quant Time: Jan 27 04:47:07 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 03:18:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	3328	0.40	ng/ul	0.00
2) Naphthalene-d8	11.27	136	10437	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	6847	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.73	188	19952m	0.40	ng/ul	0.02
16) Chrysene-d12	21.80	240	16071	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	13365m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.82	152	6863	0.40	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	20130	0.26	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.32	128	193605	6.624	ng/ul#	88
5) 2-Methylnaphthalene	12.89	142	59042	3.025	ng/ul	99
8) Acenaphthene	15.08	153	16250	0.636	ng/ul	97
9) Fluorene	16.03	166	12499	0.474	ng/ul#	72
11) Pentachlorophenol	17.36	266	545817	141.240	ng/ul#	56
12) Phenanthrene	17.75	178	58956	0.879	ng/ul#	85
13) Anthracene	17.85	178	9918	0.161	ng/ul#	91
15) Fluoranthene	19.73	202	24043	0.277	ng/ul	84
17) Pyrene	20.08	202	13703m	0.254	ng/ul	
18) Benzo(a)anthracene	21.79	228	3518	0.070	ng/ul#	88
19) Chrysene	21.84	228	2365	0.045	ng/ul	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA E\DATA\BE012618\
Data File : BE095149.D
Acq On : 27 Jan 2018 3:52
Operator : SJ/JU
Sample : J1243-01
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6B25

Manual Integrations
APPROVED

Sohil
1/29/2018 6:31:02 PM

Quant Time: Jan 27 04:47:07 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 27 03:18:43 2018
Response via : Initial Calibration

