

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
 Data File : BE095172.D
 Acq On : 27 Jan 2018 18:22
 Operator : SJ/JU
 Sample : J1220-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6A04

Quant Time: Jan 28 03:06:51 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 03:18:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	3025	0.40	ng/ul	0.00
2) Naphthalene-d8	11.27	136	9605	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	5995	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	12597	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	13650	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	14288	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.82	152	3330	0.21	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	9324	0.19	ng/ul	0.00
Target Compounds				Ovalue		
11) Pentachlorophenol	17.36	266	238	0.098	ng/ul	96
12) Phenanthrene	17.75	178	1602	0.038	ng/ul#	92
15) Fluoranthene	19.72	202	1251	0.023	ng/ul	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
Data File : BE095172.D
Acq On : 27 Jan 2018 18:22
Operator : SJ/JU
Sample : J1220-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6A04

Quant Time: Jan 28 03:06:51 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 27 03:18:43 2018
Response via : Initial Calibration

