

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
 Data File : BE095179.D
 Acq On : 27 Jan 2018 22:33
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.414

Quant Time: Jan 28 02:36:14 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 03:18:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2746	0.40	ng/ul	0.00
2) Naphthalene-d8	11.27	136	8101	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	4658	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	10427	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	12305	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	11491	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	5482	0.42	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	15851	0.40	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.32	128	9052	0.399	ng/ul#	89
5) 2-Methylnaphthalene	12.89	142	6251	0.413	ng/ul	99
7) Acenaphthylene	14.75	152	9512	0.424	ng/ul	97
8) Acenaphthene	15.08	153	7023	0.404	ng/ul	97
9) Fluorene	16.03	166	7586	0.423	ng/ul#	74
11) Pentachlorophenol	17.36	266	1088	0.539	ng/ul	96
12) Phenanthrene	17.75	178	14585	0.416	ng/ul#	87
13) Anthracene	17.85	178	14360	0.447	ng/ul#	91
15) Fluoranthene	19.73	202	17774	0.391	ng/ul	84
17) Pyrene	20.08	202	19635	0.476	ng/ul#	79
18) Benzo(a)anthracene	21.79	228	16674	0.433	ng/ul#	85
19) Chrysene	21.84	228	16019	0.401	ng/ul	96
21) Benzo(b)fluoranthene	23.62	252	14978	0.369	ng/ul	85
22) Benzo(k)fluoranthene	23.67	252	15179	0.409	ng/ul#	87
23) Benzo(a)pyrene	24.32	252	14481	0.400	ng/ul#	80
24) Indeno(1,2,3-cd)pyrene	27.24	276	16202	0.408	ng/ul#	80
25) Dibenzo(a,h)anthracene	27.27	278	13271	0.409	ng/ul#	88
26) Benzo(g,h,i)perylene	28.12	276	13468	0.401	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
Data File : BE095179.D
Acq On : 27 Jan 2018 22:33
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.414

Quant Time: Jan 28 02:36:14 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 27 03:18:43 2018
Response via : Initial Calibration

